

Good practice recommendations for the use of time-lapse technology[†]

ESHRE Working group on Time-lapse technology, Susanna Apter¹, Thomas Ebner², Thomas Freour³, Yves Guns⁴, Borut Kovacic⁵, Nathalie Le Clef⁶, Monica Marques⁷, Marcos Meseguer⁸, Debbie Montjean⁹, Ioannis Sfontouris¹⁰, Roger Sturmey¹¹, and Giovanni Coticchio^{12,*}

¹Livio, Stockholm, Sweden ²Department of Gynecology, Obstetrics, and Gynecological Endocrinology, Kepler Universitätsklinikum, Linz, Austria ³Médecine de la Reproduction, CHU de Nantes, Nantes, France ⁴Center for Reproductive Medicine, UZ Brussel, Brussels, Belgium ⁵Department of Reproductive Medicine and Gynecologic Endocrinology, Univerzitetni klinični center Maribor, Maribor, Slovenia ⁶European Society of Human Reproduction and Embryology, Grimbergen, Belgium ⁷CEMEARE, Lisbon, Portugal ⁸IVF Laboratory, Instituto Valenciano de Infertilidad, Valencia, Spain ⁹Médecine et Biologie de la Reproduction, Hôpital Saint Joseph, Marseille, France ¹⁰Assisted Reproduction Unit, Eugonia, Athens, Greece ¹¹Centre for Atherothrombosis and Metabolic Disease, Hull York Medical School, University of Hull, Hull, UK ¹²9.baby Family and Fertility Center, Bologna, Italy

*Correspondence address. 9.baby Family and Fertility Center, via Dante 15, 40125 Bologna, Italy. E-mail: giovanni.coticchio@gmail.com  <https://orcid.org/0000-0003-1635-9205>

Submitted on December 5, 2019; resubmitted on December 5, 2019; editorial decision on January 29, 2020

STUDY QUESTION: What recommendations can be provided on the approach to and use of time-lapse technology (TLT) in an IVF laboratory?

SUMMARY ANSWER: The present ESHRE document provides 11 recommendations on how to introduce TLT in the IVF laboratory.

WHAT IS KNOWN ALREADY: Studies have been published on the use of TLT in clinical embryology. However, a systematic assessment of how to approach and introduce this technology is currently missing.

STUDY DESIGN, SIZE, DURATION: A working group of members of the Steering Committee of the ESHRE Special Interest Group in Embryology and selected ESHRE members was formed in order to write recommendations on the practical aspects of TLT for the IVF laboratory.

PARTICIPANTS/MATERIALS, SETTING, METHODS: The working group included 11 members of different nationalities with internationally recognized experience in clinical embryology and basic science embryology, in addition to TLT. This document is developed according to the manual for development of ESHRE recommendations for good practice. Where possible, the statements are supported by studies retrieved from a PUBMED literature search on 'time-lapse' and ART.

MAIN RESULTS AND THE ROLE OF CHANCE: A clear clinical benefit of the use of TLT, i.e. an increase in IVF success rates, remains to be proven. Meanwhile, TLT systems are being introduced in IVF laboratories. The working group listed 11 recommendations on what to do before introducing TLT in the lab. These statements include an assessment of the pros and cons of acquiring a TLT system, selection of relevant morphokinetic parameters, selection of an appropriate TLT system with technical and customer support, development of an internal checklist and education of staff. All these aspects are explained further here, based on the current literature and expert opinion.

LIMITATIONS, REASONS FOR CAUTION: Owing to the limited evidence available, recommendations are mostly based on clinical and technical expertise. The paper provides technical advice, but leaves any decision on whether or not to use TLT to the individual centres.

WIDER IMPLICATIONS OF THE FINDINGS: This document is expected to have a significant impact on future developments of clinical embryology, considering the increasing role and impact of TLT.

STUDY FUNDING/COMPETING INTEREST(S): The meetings of the working group were funded by ESHRE. S.A. declares participation in the Nordic Embryology Academic Team with meetings sponsored by Gedeon Richter. T.E. declares to have organized workshops for Esco and receiving consulting fees from Ferring and Gynemed and speakers' fees from Esco and honorarium from Merck and MSD. T.F. received consulting fees from Vitrolife and Laboratoires Genévrier, speakers' fees from Merck Serono, Gedeon Richter, MSD and Ferring and research grants from Gedeon Richter and MSD. M.M. received sponsorship from Merck. M.M.E. received speakers' fees from Merck, Ferring and MSD.

R.S. received a research grant from ESHRE. G.C. received speakers' fees from IBSA and Excemed. The other authors declare that they have no conflict of interest.

TRIAL REGISTRATION NUMBER: N/A.

DISCLAIMER: This Good Practice Recommendations (GPR) document represents the views of ESHRE, which are the result of consensus between the relevant ESHRE stakeholders and are based on the scientific evidence available at the time of preparation.

ESHRE's GPRs should be used for information and educational purposes. They should not be interpreted as setting a standard of care or be deemed inclusive of all proper methods of care nor exclusive of other methods of care reasonably directed to obtaining the same results. They do not replace the need for application of clinical judgment to each individual presentation, nor variations based on locality and facility type.

Furthermore, ESHRE GPRs do not constitute or imply the endorsement, or favouring of any of the included technologies by ESHRE.

†ESHRE Pages content is not externally peer reviewed. The manuscript has been approved by the Executive Committee of ESHRE.

Key words: time-lapse technology / ESHRE / guideline / embryology / embryo selection / morphokinetics / ART

WHAT DOES THIS MEAN FOR PATIENTS?

In most fertility treatments (IVF and ICSI), eggs retrieved from the woman are fertilized with sperm from the man in the laboratory. These fertilized eggs are then cultured in the laboratory for a few days before being transferred as embryos to the woman's womb or frozen. During this process, embryologists regularly check the development of the embryos under the microscope to make sure they are developing well and to be able to pick the best embryo to be transferred. To have good-quality embryos and a good chance of pregnancy, it is important that the embryos are cultured in a stable environment, which means using an incubator with a fixed temperature and oxygen level. Alterations of such conditions should be kept to a minimum, but inevitably embryologists will need to take the embryos out of the incubator at least once a day to monitor their development.

Time-lapse technology (TLT) systems are used in some laboratories to facilitate embryo monitoring. In a TLT incubator, images of embryo development are recorded at regular intervals of 5–15 min. This allows the embryologists to assess embryo development thoroughly in a dynamic fashion without removing them from the incubator. Some TLT systems use specific computer programs that assist in the assessment of embryos based on changes in shape/structure occurring over time and help in the ranking of embryos depending on their developmental ability.

The TLT has been introduced in several IVF labs, and studies have reported on the clinical outcomes (pregnancy rates) after its use. Overall, the studies do not allow us to conclude that better pregnancy rates can be achieved after using a TLT system compared with standard incubation and assessment. However, using a TLT system may have other benefits, for example a better and more flexible management of laboratory workload. Drawbacks of the time-lapse technology include the need for specific training on using this type of equipment. By describing the benefits, drawbacks and impact of a TLT system, this paper provides recommendations for good practice, which will help embryologists to make decisions about the choice of a specific device and help them to use it appropriately.

Patients are sometimes offered use of the time-lapse technology (with or without extra costs). The information on the benefits, drawbacks and impact of a time-lapse system in this paper will help patients in discussing the topic at the IVF clinic and deciding on how to proceed.

Introduction

An optimal incubation environment and accurate embryo selection are two defining factors for the successful outcome of IVF treatment. During *in vitro* culture, embryos are typically assessed by morphological grading in order to predict embryo developmental competence and implantation potential. The features evaluated may include the morphology of pronuclei (PN) and nucleoli, stage-specific number and size of blastomeres, fragmentation, multinucleation, blastocyst expansion and inner cell mass (ICM) and trophectoderm (TE) appearance (Cummins *et al.*, 1986; Scott, 2003; Ahlstrom *et al.*, 2011; Fulka *et al.*, 2015; De los Santos *et al.*, 2016; Otsuki *et al.*, 2017). Traditional morphological evaluation is performed at static time points and thus provides a 'snap-shot' of embryo development. Furthermore, it usually requires physical removal of the embryos from the incubator, exposing them to fluctuations in temperature, pH and oxygen levels. Crucially, this approach has limited ability to predict embryo developmental competence and ongoing pregnancy, with high intra- and inter-observer variability (Rijnders and Jansen, 1998; Guerif *et al.*,

2007). In an attempt to standardize morphological evaluation across different laboratories, a consensus on the timings and characteristics of morphology assessment of human embryos was published by ESHRE and Alpha Scientists in Reproductive Medicine (Alpha Scientists in Reproductive Medicine and ESHRE Special Interest Group of Embryology, 2011). Although this was undoubtedly a step in the right direction, the limitations of static morphology evaluation were not overcome.

Although time-lapse technology (TLT) was introduced in ART many years ago (Payne *et al.*, 1997), it was not until 2010 that TLT shifted from a mere observation of human embryos while in culture, to a selection and prediction tool. Wong *et al.* (2010) described an algorithm able to predict blastocyst formation by day 2 of embryo culture, based on cell division timings (Wong *et al.*, 2010). The year of 2011 marked the official introduction of TLT in the embryology laboratory, when embryo implantation was shown to be associated with specific cell division timing parameters, introducing the term 'morphokinetics' (Meseguer *et al.*, 2011). The introduction of TLT has enabled both an increase in the number of observations and the dynamic assessment

of developing embryos. In parallel, TLT offers an uninterrupted culture environment, minimizing embryo handling and the need to expose embryos to conditions outside of the incubator (Meseguer *et al.*, 2012).

A TLT system typically comprises a stand-alone incubator with one or more integrated inverted microscopes coupled to a digital camera. Alternatively, and less commonly, an optical system can be placed inside a conventional incubator. In both cases, digital images are collected at regular intervals and at different focal planes throughout embryo development and subsequently processed into videos. With this information, TLT enables embryologists to record preimplantation embryo development in a dynamic, real-time manner and permits the interpretation of morphokinetic events more precisely. Data from these observations can be annotated and analysed using integrated TLT software, facilitating the development of more complex embryo selection/deselection algorithms (Ciray *et al.*, 2014, Rubio *et al.*, 2014).

This paper will describe different types of TLT systems, discuss the potential benefits and uses of TLT and evaluate the impact on laboratory workflow, in order to inform IVF clinics as they choose a system appropriate for their own needs. This paper is not intended as a manual on the use of TLT, nor does it provide a systematic description of clinical evidence. A meta-analysis of randomized controlled trials (RCT) assessing clinical outcomes after TLT was recently published (Armstrong *et al.*, 2019).

Materials and Methods

ESHRE good practice recommendations are developed based on the *Manual for development of recommendations for good practice* (Vermeulen *et al.*, 2018), which can be consulted at the ESHRE website (www.eshre.eu/guidelines). The manual describes a nine-step procedure for writing recommendations documents by the working group supported by the ESHRE methodological expert.

The current paper is the result of a 2-day consensus meeting and three online meetings of the working group. In preparation of the consensus meeting, information was collected by means of published surveys, manufacturer information and narrative reviews. In addition, relevant published data were collected from a literature search. We searched PUBMED from insertion to 23 January 2019 combining search terms (including MESH terms and synonyms) for ART/IVF and time-lapse. Papers not in English or not focused on TLT for ART were excluded. All other references were assessed, and relevant papers selected. Each working group member prepared a draft of a pre-allocated section, after which these were discussed until consensus within the group was reached. After the meeting, all ESHRE members were invited to submit comments during stakeholder review of the draft; it was published on the ESHRE website between 21 June and 2 August 2019. Fourteen people participated in the stakeholder review and submitted comments (Supplementary Data). Each comment was documented in a review report, and appropriate changes were made in the manuscript. A review report is published on the ESHRE website (www.eshre.eu/guidelines).

Recommendations

A list of recommendations for clinics before getting started with TLT is provided below.

List of recommendations for clinics before getting started with time-lapse technology in human IVF

- Clearly identify the reasons to introduce a TLT system
- Assess pros and cons of acquiring a TLT system, both financially and operatively
- Identify whether morphokinetic parameters will be used in selection/deselection /ranking of embryos
- Identify (from scientific literature) the morphokinetic parameters of interest and assess how to monitor and use them
- Find the suitable system based on considerations of culture conditions/systems and other costs, including hardware maintenance and software upgrades
- Evaluate technical/customer support available, including accessibility, the level of embryologist support and the expertise the manufacturer will provide to your team
- Seek appropriate installation and training from the manufacturer/distributor
- Develop an internal checklist, based on a user requirement specification for the system, identifying and matching what the clinic/laboratory wants in a system e.g. type of gas, humidity, footprint, capacity, type of dish, software, cost, supply chain and manufacturer support.
- Once introduced in the lab, find the appropriate system settings
- Identify and train one or more embryologists (depending on the size of the laboratory) who will develop the role of "TLT referent"; the designated person(s) will be responsible for the annotation of morphokinetic variables (to avoid initially inter-operator variations with other members of staff) and for the implementation of quality control programs
- Educate clinic staff on the current evidence behind TLT in order to counsel patients alongside offering the technology.

TLT: time-lapse technology.

Why clinics can use TLT (significance of TLT) for embryo assessment

The identification of the embryo with the best prognosis remains an unmet need in IVF. This section will evaluate whether morphokinetic embryo assessment by TLT may assist in achieving this goal.

Embryo assessment based on fertilisation markers

Markers of embryo quality at early stages of development are of particular value to clinics where extended embryo culture is not feasible. Following pioneering research by Payne *et al.*, TLT enabled Coticchio and colleagues to draw an in-depth map of events occurring during fertilisation, which may be putative indicators of embryo quality (Payne *et al.*, 1997; Coticchio *et al.*, 2018). Twenty-eight parameters that were previously unknown or were poorly documented were described. The time intervals between four morphokinetic events were shown to predict embryo quality on day 3. These were cytoplasmic halo

appearance → disappearance; halo appearance → PN fading; PN fading → first cleavage (t2); and (iv) male PN appearance → male PN fading (tPNf) (Coticchio *et al.*, 2018). Further studies assessing these markers as predictors of embryo quality on day 5 and clinical outcome are required, but TLT is the only existing technology that enables assessing embryos based on such criteria. Although time of polar body emission (tPB2) and PN morphology did not predict live birth, tPNf was associated with live birth, i.e. the tPNf of zygotes resulting in live birth was significantly longer than the tPNf of the no live birth group (Azzarello *et al.*, 2012). Furthermore, it was reported that erratic PN movement within the cytoplasm and delayed fading of nuclear envelopes are indicative of compromised embryo developmental potential (Athayde Wirka *et al.*, 2014).

Embryo assessment and cleavage features

Discrete cleavage anomalies (Table I), mostly undetectable with static embryo assessment, have been described and correlated with embryo quality, chromosomal status and implantation potential. Future studies are still needed to standardize, sub-categorize and more clearly define irregularly cleaved embryos (Lagalla *et al.*, 2017).

Wong and colleagues have shown that blastocyst development can be predicted with high sensitivity (94%) and specificity (93%) based on parameters identified by tracking an embryo up to the four-cell stage, namely, the time interval between the end of the first mitosis and the initiation of the second (duration of two-cell stage) and the time interval between the second and third mitoses (duration of the three-cell stage) (Wong *et al.*, 2010).

Guidelines were proposed on the nomenclature and annotation of the events observed during embryo development followed with a TLT system (Ciray *et al.*, 2014). The variable and the description of the events and intervals are summarized in Table II.

Five-cell cleavage timing and intervals during two cleavages (t5 and s2, cc2) (Table II) were shown to be the most predictive parameters for embryo viability and implantation (Wong *et al.*, 2010; Meseguer *et al.*, 2011). Recently, an association between irregular division (Liu *et al.*, 2014a; Desai *et al.*, 2018b), start time of blastulation (tSB), expansion (tEB), the interval tEB-tSB and both ploidy and aneuploidy status with odds of live birth was reported (Desai *et al.*, 2018b, Fishel *et al.*, 2018).

Since the routine introduction of TLT, there have been numerous attempts to assess the clinical and biological significance of the parameters described in Table II. A non-exhaustive summary of these studies is available in Table III. However, it is difficult to compare the outcomes of these different studies since the methodologies used are not consistent.

TLT and ploidy status

Embryo ploidy status is probably the most critical factor impacting an embryo's implantation potential. PGT for aneuploidy (PGT-A) has greatly improved over the last few years and allows the accurate evaluation of embryo chromosomal status. However, PGT-A is not permitted in some countries, and there remains some debate regarding its cost-effectiveness and/or clinical relevance (Dahdouh *et al.*, 2015; Sermon *et al.*, 2016; Griffin and Ogur, 2018; Neal *et al.*, 2018; Penzias *et al.*, 2018; Rosenwaks *et al.*, 2018; Lee *et al.*, 2019; Somigliana *et al.*, 2019). As TLT provides extensive information on embryo development *in vitro*, it is postulated that morphokinetic parameters

could be associated with embryo ploidy, thus providing a cheaper, faster and less invasive method for the evaluation of embryo ploidy status than PGT-A (Chavez *et al.*, 2012; Campbell *et al.*, 2013). A comprehensive review of the literature on the predictive value of morphokinetic parameters for embryo ploidy status was reported recently (Reignier *et al.*, 2018). A total of 13 studies were included, which had significant heterogeneity in terms of design, inclusion criteria, embryo biopsy, statistical approach and outcome measures. While most studies found significant differences in morphokinetic parameters between euploid and aneuploid embryos, none provided evidence sufficient to recommend the clinical use of TLT for embryo ploidy assessment. The same conclusion was reached in another contemporary review where the association between morphokinetics and aneuploidy was discussed (Zaninovic *et al.*, 2017). However, the combination of PGT-A with morphokinetic analysis may help in selecting the embryo with the highest implantation potential (Rocafort *et al.*, 2018).

Training/teaching

TLT provides an excellent tool for teaching embryology and standardizing assessment. Since embryos can be examined without removal from the incubator to assess their morphology and dynamic events, the time factor is no longer an issue and detailed assessment is also feasible *a posteriori*. Visual examples of standard morphology assessment and examples of normal and unusual, probably abnormal, cleavage patterns can easily be stored and used as learning material.

Officially recognized training programmes to direct staff in the use of TLT devices and morphokinetic annotation remain to be developed. Such programmes should enable a thorough understanding of the technical and theoretical principles governing equipment operation; acquisition of manual skills to set up and maintain the embryo culture conditions required by the device; and attainment of competences relevant to morphokinetic annotation and cycle treatment data input.

Quality control

Intra- and inter-observer variability impacts on static morphological embryo scoring and evaluation of morphokinetic criteria (Sundvall *et al.*, 2013). Several factors can affect precision and reproducibility of morphokinetic annotation by TLT; examples include the selection of an appropriate focal plane for the observation of spatially restricted events, consensus on when to annotate events that are occurring gradually (e.g. pronuclear formation or compaction) and mere definition of the parameters of interest. Initial experiences aimed at assessing intra- and inter-operator variability in annotation of morphokinetic parameters were reassuring (Sundvall *et al.*, 2013). Overall, inter-observer annotation, subject to possible biases due to assessment of morphokinetic behaviours not amenable to precise quantitative measurement was found to have an almost perfect agreement, although the degree of conformity was not the same for the diverse parameters. For example, the measurements with the highest degree of agreement were those relevant to pronuclear fading, nuclear appearance and disappearance at the two-cell stage and achievement of full blastocyst hatching. On the other hand, parameters that were less consistently annotated included pronuclear appearance, multinucleation, blastomere evenness and number of collapses during blastocyst expansion. In general, intra-observer annotations (typically subject to random errors) were characterized by an even higher, although not

Table I Atypical human embryo cleavage features observed with time-lapse technology versus classic embryo morphology assessment once per day.

Name of feature	Explanation	Observed exclusively or better by TLT	References
Abnormal syngamy	Erratic PN movement in the cytoplasm	Exclusively	(Athayde Wirka <i>et al.</i> , 2014, Coticchio <i>et al.</i> , 2018)
Asynchronous appearance of two pronuclei	Disappearance of one and appearance of another pronucleus	Exclusively	(Coticchio <i>et al.</i> , 2018)
Differently sized pronuclei	Difference in pronuclear areas immediately before pronuclear membrane fading	Exclusively	(Otsuki <i>et al.</i> , 2017)
Pronuclei reappearance	Pronuclei fading and reappearance	Exclusively	(Coticchio <i>et al.</i> , 2018)
Aberrant behaviour of female pronucleus	Extrusion of the third PB instead of female pronucleus formation	Exclusively	(Mio <i>et al.</i> , 2014)
Fragmentation of pronuclei	Formation of micronuclei	Better	(Mio <i>et al.</i> , 2014)
Fusion of pronuclei	A pronucleus formed by the fusion of two preexisting pronuclei	Exclusively	(Mio <i>et al.</i> , 2014)
Unipolar cleavage furrow	Appearance of cleavage furrow on one site of the zygote	Exclusively	(Hojnik <i>et al.</i> , 2016, Wong <i>et al.</i> , 2010)
Tripolar cleavage furrow	Appearance of three cleavage furrows on the zygote	Exclusively	(Wong <i>et al.</i> , 2010)
Pseudofurrows	Zygote presenting oolemma ruffling before cytokinesis	Exclusively	(Athayde Wirka <i>et al.</i> , 2014, Wong <i>et al.</i> , 2010)
Absent cleavage	Arrest in zygote stage despite normal fertilisation	Better	(Barrie <i>et al.</i> , 2017)
Direct cleavage	Cleavage of zygote to three cells (trichotomous mitosis) or one blastomere to three cells in the first ($t_3 - t_2 = 0$) or second cell division cycle (two cells to five or six cells), but this should be distinguished from rapid cleavage ($t_3 - t_2 < 5$ h)	Exclusively	(Athayde Wirka <i>et al.</i> , 2014, Barrie <i>et al.</i> , 2017, Fan <i>et al.</i> , 2016, Lagalla <i>et al.</i> , 2017, Meseguer <i>et al.</i> , 2011, Rubio <i>et al.</i> , 2012, Zhan <i>et al.</i> , 2016)
Reverse cleavage	Fusion of two cells into one blastomere	Exclusively	(Barrie <i>et al.</i> , 2017, Desai <i>et al.</i> , 2014, Goodman <i>et al.</i> , 2016, Liu <i>et al.</i> , 2014b)
Blastomere movement	Prolonged blastomere movement induced by delay in pronuclear fading and first cell division	Exclusively	(Ezoe <i>et al.</i> , 2019)
Multinucleation	Blastomere with > 1 nucleus	Better	(Balakier <i>et al.</i> , 2016, Desai <i>et al.</i> , 2014, Ergin <i>et al.</i> , 2014, Goodman <i>et al.</i> , 2016, Hashimoto <i>et al.</i> , 2016)
Internalization of cellular fragments	Fragments reabsorbed into one mother blastomere	Exclusively	(Hardarson <i>et al.</i> , 2002, Mio <i>et al.</i> , 2014)
Irregular chaotic division	Disordered cleavage behaviour with uneven cleavages and fragmentation	Better	(Athayde Wirka <i>et al.</i> , 2014, Barrie <i>et al.</i> , 2017)
Early compaction	Formation of tight junctions between blastomeres in day 3 or even day 2 embryos	Better	(Iwata <i>et al.</i> , 2014, Le Cruguel <i>et al.</i> , 2013)
Cell exclusion	Exclusion of one or more blastomeres from the formation of compact morula or blastocyst	Better	(Coticchio <i>et al.</i> , 2019, Lagalla <i>et al.</i> , 2017)
Blastocyst collapse	Complete or almost complete disappearance of blastocoel and consequent blastocyst shrinkage	Better	(Bodri <i>et al.</i> , 2016b, Kovacic <i>et al.</i> , 2018, Marcos <i>et al.</i> , 2015)

TLT: time-lapse technology

statistically significant, coefficient of consistency. Interestingly, in this class, the degree of agreement of each parameter reflected the same trend reported for the inter-observer comparisons (as indicated by * in

Table II). Good intra- and inter-observer agreement was also reported in more recent studies (Adolfsson and Andershed, 2018; Storr *et al.*, 2018).

Table II Nomenclature of morphokinetics parameters.

	Terminology	Description of the event
	tPB2	The second polar body is completely detached from the oolemma
	tPNa	Appearance of individual pronuclei; tPN1a, tPN2a, tPN3a, ...
	tPNf*	Time frame of pronuclei fading; tPN1f; tPN2f...
	tZ	Time of PN scoring (last time frame before tPNf)
	tn*	First time frame at which an embryo reaches <i>n</i> number of blastomeres (e.g. t2, t3, t4)
	tTM	Trichotomous mitosis at different stages
	tSC	First evidence of compaction
	tM	Time of completion of compaction process (in case some blastomeres are excluded, it might be difficult to assess the real time frame)
	tSB	Initiation of blastulation (first frame in which the blastocoel is visible)
	tB	Full blastocyst (last frame before zona starts to thin)
	tE or tEB	Initiation of expansion; first frame of zona thinning (also called TEyB 'y' corresponds to morphology of inner cell mass; 'z' corresponds to morphology of trophoctoderm cells)
	tHN	Herniation; end of expansion phase and initiation of hatching process (also called tHNyz)
	tHD or tHB*	Fully hatched blastocyst (also called tHDyz)
Dynamic events and time intervals	Psyn	Syngamy, time from PN fading to the first cytokinesis
	Not mentioned	Time between nuclear envelope breakdown and subsequent division to two cells
	s2	Time between division to three cells and subsequent division to four cells
	s3	Time between division to five cells and subsequent division to eight cells
	ECC1	Duration of the first cell cycle (t2-tPB2)
	cc2	Blastomere cell cycle: Duration of the second cell cycle ($a = t3 - t2$, $b = t4 - t2$)
	cc3	Blastomere cell cycle: Duration of the third cell cycle ($a = t5 - t4$, $b = t6 - t4$, $c = t7 - t4$, $d = t8 - t4$)
	ECC2	Embryo cell cycle: t4-t2
	ECC3	Embryo cell cycle: t8-t4
	Blastocyst contraction	A decrease in blastocoel volume
Cryopreserved/ warmed blastocyst	tRE	Time of the start of re-expansion (first frame in which the blastocoel reforms or increases in size)
	tCRE	Time of completion of re-expansion (first frame the blastocyst occupies the whole perivitelline space)

General comment: depending on the configuration of the TLT system, some events may not be seen. Table adapted from consensus paper (Ciray *et al.*, 2014). Time zero (t0) may change from one study to another (mid-time for ICSI, standard IVF insemination, tPB2 or tPNf). These inconsistencies have to be taken into account when comparing data from different studies (Kaser and Racowsky, 2014).

*Parameters with the highest concordance between operators.

Clearly, technical differences between different TLT devices may also limit annotation consistency. For example, TLT devices may differ in time intervals between two consecutive image acquisitions, number of focal planes (Z resolution) or in the quality of images collected. These differences may have implications for events occurring rapidly (e.g. pronuclear fading) or for morphological characteristics requiring precise description (e.g. arrangement of nuclear precursor bodies). Nevertheless, a comparison of two different TLT devices showed that inter-laboratory variability clusters mostly at two specific developmental intervals, one delimited by extrusion of the second PB and pronuclear formation, and another spanning the eight-cell and the morula stages (Martinez-Granados *et al.*, 2017). Overall,

inter-laboratory agreement between different TLT devices was high, although it was similar or lower compared with conventional morphological observation, depending on the equipment used (Martinez-Granados *et al.*, 2017). Taken together, these experiences are important in order to assess the reliability of the TLT approach, but they cannot be considered conclusive, and call for more extensive analyses.

At present, automated annotation has not solved the question of fidelity of morphokinetic analysis. Automation requires human supervision to correct possible, but recurrent, annotation inaccuracies that may affect the performance of prediction models for embryo selection. Therefore, similar to other activities, each laboratory should

Table III Parameters with biological/clinical significance.

Markers	Prediction/outcome	Reference
Time interval cytoplasmic halo appearance → disappearance	Embryo quality on day 3	(Coticchio <i>et al.</i> , 2018)
Time interval halo appearance → PN fading	Embryo quality on day 3	(Coticchio <i>et al.</i> , 2018)
Time interval PN fading → first cleavage (t2)	Embryo quality on day 3	(Coticchio <i>et al.</i> , 2018)
Time interval male PN appearance → male PN fading	Embryo quality on day 3	(Coticchio <i>et al.</i> , 2018)
PNs movement and fading	Blastocyst formation	(Athayde Wirka <i>et al.</i> , 2014)
Appearance of nuclei after first cleavage	Pregnancy success	(Lemmen <i>et al.</i> , 2008)
Duration of the first cytokinesis	Blastocyst formation	(Wong <i>et al.</i> , 2010)
Time interval between the end of the first mitosis and the initiation of the second	Blastocyst formation	(Wong <i>et al.</i> , 2010)
Time interval between the second and third mitoses	Blastocyst formation	(Wong <i>et al.</i> , 2010)
tPNf	Live birth	(Azzarello <i>et al.</i> , 2012)
	Implantation	(Aguilar <i>et al.</i> , 2014, Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c, Wu <i>et al.</i> , 2016)
tPB2	Implantation	(Aguilar <i>et al.</i> , 2014)
Length of s-phase	Implantation	(Aguilar <i>et al.</i> , 2014)
	Implantation	(Meseguer <i>et al.</i> , 2011, Mizobe <i>et al.</i> , 2016a, Wu <i>et al.</i> , 2016)
t2	Blastocyst formation	(Mizobe <i>et al.</i> , 2018)
	Top-quality blastocyst formation	(Mizobe <i>et al.</i> , 2016a)
	Embryo quality on day 3	(Coticchio <i>et al.</i> , 2018)
	Implantation	(Meseguer <i>et al.</i> , 2011)
t3	Implantation	(Carrasco <i>et al.</i> , 2017, Freour <i>et al.</i> , 2013, Meseguer <i>et al.</i> , 2011, Mizobe <i>et al.</i> , 2016a, Wu <i>et al.</i> , 2016)
	Top-quality blastocyst formation	(Mizobe <i>et al.</i> , 2016a)
	Implantation	(Meseguer <i>et al.</i> , 2011)
t4	Top-quality blastocyst formation	(Mizobe <i>et al.</i> , 2016a)
	Implantation	(Meseguer <i>et al.</i> , 2011)
t5	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
t6	Implantation	(Carrasco <i>et al.</i> , 2017)
	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
t7	Implantation	(Dal Canto <i>et al.</i> , 2012)
	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
t8	Implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c)
	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
tn	Implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c)

Continued

Table III Continued.

Markers	Prediction/outcome	Reference
Mean duration of two-cell stage	Implantation	(Meseguer <i>et al.</i> , 2011) (Rubio <i>et al.</i> , 2012)
	Expanded blastocyst formation	(Dal Canto <i>et al.</i> , 2012)
	Blastocyst development	(Conaghan <i>et al.</i> , 2013, Cruz <i>et al.</i> , 2012, Wong <i>et al.</i> , 2010)
Mean duration of three-cell stage	Implantation	(Meseguer <i>et al.</i> , 2011)
	Blastocyst development	(Conaghan <i>et al.</i> , 2013, Cruz <i>et al.</i> , 2012, Wong <i>et al.</i> , 2010)
	Expanded blastocyst formation	(Dal Canto <i>et al.</i> , 2012)
tM	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
	Blastocyst formation and implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c, Motato <i>et al.</i> , 2016)
	No difference in implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c)
tSC	Implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c)
tSB	Top-quality blastocyst formation	(Fishel <i>et al.</i> , 2018, Storr <i>et al.</i> , 2015)
	Implantation	(Goodman <i>et al.</i> , 2016, Mizobe <i>et al.</i> , 2017)
tB	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
	Implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c)
tEB	Blastocyst formation and implantation	(Motato <i>et al.</i> , 2016)
	Implantation	(Chamayou <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2013c)
s3	Blastocyst formation	(Cetinkaya <i>et al.</i> , 2015)
	Top-quality blastocyst formation	(Storr <i>et al.</i> , 2015)
	Blastocyst formation and implantation	(Motato <i>et al.</i> , 2016)
	Implantation	(Carrasco <i>et al.</i> , 2017, Chamayou <i>et al.</i> , 2013)
cc3	Implantation	(Chamayou <i>et al.</i> , 2013)
Blastocyst contraction	Implantation rate	(Marcos <i>et al.</i> , 2015, Vinals Gonzalez <i>et al.</i> , 2018)
tRE, tCRE	Pregnancy	(Ebner <i>et al.</i> , 2017)
Post thawing blastocyst re-expansion speed (tCRE-tRE)	Pregnancy and pregnancy loss	(Ebner <i>et al.</i> , 2017)
	Live birth	(Kovacic <i>et al.</i> , 2018)

implement appropriate programmes of quality control and assurance (De los Santos *et al.*, 2016).

On a different level, TLT has significant relevance for other laboratory activities. For example, differences in embryo morphokinetics, as revealed by TLT, may be valuable endpoints against which to compare consumables, cryopreservation protocols and devices introduced in the IVF laboratory (Ferrick *et al.*, 2019). TLT also offers the opportunity to sharpen the sensitivity of mouse embryo assays. Wolff *et al.* (2013) reported that a morphokinetic algorithm was able to detect alterations in mouse embryo development caused by media contaminants and lots of toxic mineral oil, while the same culture conditions did not affect blastocyst rate (Wolff *et al.*, 2013). Deviant morphokinetic patterns can therefore represent an early warning of altered culture conditions.

Implications of TLT

Impact on embryo culture conditions

Culture medium

Embryos *in vitro* are exposed to numerous physical and chemical stressors (Wale and Gardner, 2016), which creates an environment that can impact on the developing embryo. Amongst these external factors, the culture medium used is a crucial one. Improvements in culture conditions have come primarily from modifications in media formulations that have been developed according to two doctrines. On the one hand, there is the attempt to satisfy the perceived changing requirements of the human embryo in a manner that is analogous to the environmental changes it would encounter as it would move *in vivo* from the oviduct to the uterus (Barnes *et al.*, 1995). The approach to address this concept is to fine-tune media composition in order to fulfil the needs of the embryo—so called ‘sequential media’. On the other hand, it has been hypothesized that it is of benefit to supply all nutrients, and the embryo will metabolize them according to its demand—so called ‘single-step media’ (Summers *et al.*, 1995). Results from studies in conventional incubators remain inconclusive as to whether one culture system is superior to the other (Sepulveda *et al.*, 2009; Sfontouris *et al.*, 2016; Werner *et al.*, 2016).

There arises a question of whether the increased resolution of TLT might identify more subtle differences, e.g. in morphokinetic behaviour, between sequential and single-step media. Ciray *et al.* were the first to compare the two approaches to embryo culture using TLT. Randomisation of mature oocytes was carried out and followed by ICSI. On day 3 of culture, those embryos in sequential culture had their medium replaced whereas the single step group had their culture medium replenished with a fresh infusion of the same medium (Ciray *et al.*, 2012). The authors found that in single-step medium, fading of PN (tPNf) and cleavage up to five-cell stage (t2 → t5) took place significantly earlier compared to counterparts grown in sequential medium. In implanted embryos, t2 and t4 were significantly shorter with a single-step medium (Ciray *et al.*, 2012). Recently, these data were, at least in part, confirmed by Kazdar *et al.*, who reported an accelerated first mitotic cell cycle (tPNf → t2) with a single-step medium (Kazdar *et al.*, 2017). In contrast, others have been unable to identify morphokinetic differences between embryos grown in sequential or single-step culture (Basile *et al.*, 2013; Schiewe *et al.*, 2018). However, it is possible that any developmental delay at earlier times may be compensated at later stages. Indeed, in a

recent multicentre trial, culture in a single-step medium designed specifically for TLT resulted in a longer t7 and t8, but by blastulation (tSB) the differences were no longer present (Hardarson *et al.*, 2015).

Crucially, no study has yet demonstrated any effect of single-step or sequential media on implantation and pregnancy rates. The uninterrupted culture, which avoids the need for media replenishment and thus minimizing culture disruption and stress to the embryos, may be preferred for practical reasons. However, renewing media on day 3 does neither influence morphokinetics nor implantation and live birth (Costa-Borges *et al.*, 2016).

Thus, data to date have been unable to demonstrate conclusive superiority of either single-step nor sequential media in terms of clinical outcomes when used in conjunction with TLT incubators.

Oxygen tension

It is now widely accepted that the oxygen tension of the mammalian female reproductive tract is between 2 and 8% (Fischer and Bavister, 1993). Exposure of embryos to atmospheric oxygen tension is associated with a higher production of reactive oxygen species (ROS) (Yang *et al.*, 1998) and may also alter gene expression (Rinaudo *et al.*, 2006), DNA methylation (Li *et al.*, 2016) and embryo metabolism (Wale and Gardner, 2012). There is evidence that embryo culture in 5% rather than ambient oxygen leads to improved pregnancy and life-birth (Meintjes *et al.*, 2009; Kovacic *et al.*, 2010; Bontekoe *et al.*, 2012). Such benefits of lower O₂ levels will almost certainly apply to TLT incubators (and as such is recommended by the supplier).

To address this, Wale and Gardner (2010) cultured murine embryos in low (5%) or high (20%) O₂ concentrations for the first 2 days, followed by culture in the same or reciprocal O₂ concentrations for a further 2 days. They reported irreversible and detrimental effects of atmospheric oxygen on mouse embryo development from the first mitosis (Wale and Gardner, 2010). More importantly, the delay in the timing of cleavages was found to be cumulative, since it became more pronounced as embryo development progressed. In addition, blastocysts that were exposed to atmospheric O₂ at any stage had significantly fewer cells compared with the 5% O₂ counterparts. In human, Kirkegaard *et al.* (2013a) found that timing of the third cleavage cycle (t5–t8) was faster for embryos cultured in 5% compared with embryos cultured in 20% O₂. However, no differences were observed in timing of the early and full blastocyst stages (Kirkegaard *et al.*, 2013a). Since the delayed development after culture in ambient O₂ was seen in the precompaction embryo only, it seems that in human the negative influence of high oxygen may be stage-specific.

Embryo density

Human embryos are capable of *in vitro* development whether cultured in groups or individually, while the embryos of many other mammals require culture in groups. For instance, contrary to grouped embryos, mouse embryos cultured individually are more sensitive to the stress caused by atmospheric O₂ (Kelley and Gardner, 2016). It is speculated that grouping such embryos may lower local O₂ concentrations and, as a consequence, reduce ROS (Wale and Gardner, 2010). In addition, embryotrophic factors may play a role in the better performance of group culture (O'Neill, 2008; Ebner *et al.*, 2010).

Kelley and Gardner (2016) were the first to use the time-lapse technique to measure the influence of embryo density on cleavage

behaviour. Although detectable from t2 (20% O₂) and t3 (5% O₂), the significant delay in individual culture culminated at the eight-cell stage (5% O₂, 1.29 h) or blastocyst stage (20% O₂, 4.76 h) (Kelley and Gardner, 2016). In a follow-up study, it was shown that embryos that had individual culture—for the entire duration of culture or any portion thereof—had fewer cells at blastocyst stage compared with those cultured in groups. This was especially notable in the ICM (Kelley and Gardner, 2017).

It is important to stress that with current TLT systems the ideal group culture is not possible due to the design of the commercially available culture dishes. There are two types of dishes, which have either multiple microwells under one drop of media, or single wells which require separate drops of media (both under mineral oil). There is evidence that the multiple microwell type better supports embryo development compared with single culture in individual drops (Chung et al., 2015). A similar effect can be achieved by simply increasing the volume of individual droplets so that they have contact with each other, but this may not be in compliance with some of the manufacturer's recommendations for dish preparation. Importantly, with the current dimensions of the dishes, and particularly the distance between microwells, any potential paracrine action of embryotrophic factors is very unlikely (Gopichandran and Leese, 2006; Ebner et al., 2010).

In a mouse model, Swain and co-workers (2012) emphasized the importance of drop size in maintaining osmolality of culture media (Swain et al., 2012). They found that using a larger volume of medium (40 µl) resulted in a significantly smaller increase of osmolality (e.g. 12 mOsm/kg) as compared to 10- and 20-µl drops. Using dishes specifically designed for time-lapse imaging, Kelley and Gardner (2017) reported that in volumes of 2 and 20 µl, only a minor increase of 4–5 mOsm/kg in osmolality was observed, which had no effect on further growth (Kelley and Gardner, 2017). This negligible shift could be caused by absorption of water by the mineral oil overlay or due to the manipulation during sampling and measuring (Heo et al., 2007).

Although the optimal osmolality for human embryo culture is still unclear, for physiological reasons most of the culture media today are specified to fall in a range of 270–290 mOsm (Sunde et al., 2016) or even lower (Baltz, 2012). However, it would be expected that *in vitro* culture of embryos will be performed within their range of osmotic tolerance (Wale and Gardner, 2016). In terms of hyperosmolality, embryo osmotic tolerance is up to 320 mOsm or even higher (e.g. 350 mOsm) considering the fact that current culture media contain strong osmolytes such as the amino acid glycine (Baltz and Tartia, 2010).

A humid atmosphere, as shown in non-TLT incubators (Fawzy et al., 2017), could counteract potential adverse effects since it may reduce the effect of fluctuations in osmolality (Yumoto et al., 2019). This stabilisation effect was strongly related to the drop volume and the mineral oil layer used (Yumoto et al., 2019). Furthermore, any theoretical drawback strongly depends on the starting osmolality of the culture medium, the length of *in vitro* culture (oocytes are most sensitive to hypo- or hyperosmolality) and whether the medium is changed on day 3 or not. It should be stressed, however, that the gradual changes in osmolality reported above never reached critical values (Swain et al., 2012; Kelley and Gardner, 2017; Yumoto et al., 2019). Using mineral oil which has been preequilibrated in a humid incubator overnight would further reduce osmotic stress (Yumoto et al., 2019).

To summarize, the current method of culturing embryos for 5–6 days in medium-sized drops of single-step or sequential media covered with mineral oil does not appear to affect osmolality and, as a consequence, development of the embryos. It is, however, strongly recommended to work with reduced oxygen.

Management of staff time, work-flow, staff training

A key strategic decision associated with investing in TLT is deciding how to implement the technology. No matter which approach is chosen, a TLT system will have a significant impact on the logistics of the laboratory. TLT eliminates the necessity of assessing embryos at fixed time points (Alpha Scientists in Reproductive Medicine and ESHRE Special Interest Group of Embryology, 2011), instead providing the flexibility of reviewing the developmental history at any appropriate time, possibly even remotely from the laboratory. This flexibility can improve efficiency as it allows better planning and timing of specific tasks (i.e. fertilisation check, embryo biopsy) and use of equipment (such as inverted microscopes). Importantly, instead of basing clinical decisions on single and static assessments, more information is available with TLT for ranking and selecting embryos. When initially implemented, staff members usually want to spend a lot of time looking at the videos generated by TLT. They will learn a lot about early embryonic development, and many questions will emerge surrounding the significance, sequence, relative timing, duration and relative importance of morphokinetic heterogeneity. It may be wise to proactively develop strategies to ensure the availability of sufficient resources during the introduction and training of staff and to manage any effect on laboratory productivity. However, once accustomed to the technology, staff members will become more efficient at making annotations.

Since TLT does not require physical removal of the embryos from the incubator, staff members can perform a more thorough assessment. The possibility to 'scroll back and forward' allows users to review the continuum of development, which should make the assessments more reliable. Moreover, the availability of the videos makes it easier to ask colleagues for a second opinion. Thus, when choosing embryos for transfer/cryopreservation, laboratories with TLT will be able to implement their deselection or ranking strategy more confidently and incisively. Laboratories may opt to only annotate in detail the morphologically good- and fair-quality embryos on the day of transfer and to give a simple morphological score for any remaining poor-quality embryos.

Policy

In most cases, laboratories are not exclusively equipped with TLT systems. When implementing a TLT approach, it is essential that clinics perform a detailed analysis to develop a tailored policy for its use with reference to their patient population. In doing so, clinics would be wise to consider a range of factors including, but not limited to, the number of units available, patient characteristics, their medical histories, the number of embryos available, day of transfer and enrolment in a PGT programme.

Staff training

Laboratories need to have appropriate standard operating procedures for tissue culture and, where used, assessment of embryos using TLT. In addition, an appropriate training programme for staff members

should be implemented, as part of a quality control programme. The training programme must clearly contain information on how to operate the TLT system and include an evaluation of the relevant time-lapse parameters. Importantly, some time-lapse parameters appear to be more difficult to assess with high consistency (Sundvall *et al.*, 2013). Therefore, any TLT training programme should be complemented with training in static morphological assessment.

It is important to inform medical and nursing staff of the new routines concerning assessment and culture. TLT can not only be used to increase understanding of embryo development but also as an important aid in making embryo assessments more descriptive, hence facilitating exchange of information amongst operators with different roles forming the IVF team.

How to introduce TLT

Different TLT systems

Currently, there are several commercially available TLT Systems. The choice of the system can be based on practical considerations, such as the laboratory workload, dimensions and the budget, or on the specifications of the individual systems. The key features of systems currently commercially available are summarized in Table IV.

As outlined in above, all TLT systems currently available require a specific culture dish, supplied by the manufacturer. Most of the culture dishes are designed for single embryo culture, for image analysis and traceability purposes. However, some of the culture dishes permit the sharing of culture media between compartments, in theory allowing exchange of soluble components, and are described by manufacturers as group culture. This may represent an important consideration when choosing a TLT system.

In addition, factors influencing a decision might include the nature of the computer software used for visualisation and analysis, and the options for annotation, which may be manual, guided or automated. A guided annotation may minimize the time spent on annotations. Furthermore, some companies offer predictive algorithms (Conaghan *et al.*, 2013; Petersen *et al.*, 2016) to be used on their equipment, which may incur additional costs. Nevertheless, it is important for each clinics to independently validate their own approach for embryo selection (see below).

Safety

Installation

Introducing a TLT incubator in the laboratory should start with the installation performed in accordance with manufacturer's instructions and should be accompanied by operational and performance qualification. As with any incubation system, TLT requires a connection to an external monitoring/alarm system, which must be tested prior to clinical use (De los Santos *et al.*, 2016). Some TLT systems allow remote follow-up of system performance. In case of emergency, troubleshooting protocols should be in place and system redundancy is required to allow, if necessary, for culture dishes to be removed and transferred to other available incubators.

Incubator

Light source. There is evidence suggesting a negative effect of light exposure on embryo development. Light emitted at 400–500 nm (blue

light) appears to be more harmful than longer wavelengths (green, orange, red light) of visible light, resulting in oxidative stress (Ottosen *et al.*, 2007). Umaoka *et al.* (1992) reported a significant reduction in the rate of first cleavage in hamster zygotes when exposed to <500 nm [blue] light for 30 min (Umaoka *et al.*, 1992). These data were confirmed in a more sensitive hamster model, also showing that light emitted at 400–500 nm resulted in a decrease in blastocyst formation and reduced blastocyst quality with increasing ICM & TE cell apoptosis. However, the detrimental effects of visible light are not only related to the spectral composition of the light, but also to the intensity and exposure time (Oh *et al.*, 2007).

Exposure frequency and duration. In a TLT incubator, an embryo may be subjected to light exposure up to 1500 times. However, even in older systems, exposing embryos approximately 300 times to white light of 80-ms exposure times does not significantly affect the fertilisation rate of ICSI, the cleavage rate or the morphological grade of embryos compared to conventional embryo scoring (Nakahara *et al.*, 2010). This suggests that there is little effect, if any, of light exposure on embryos from exposure during time-lapse observations. Intuitively, it is expected that in TLT incubators, embryos are more exposed to light. However, scalar irradiance and therefore light exposure in TLT systems are lower than with conventional morphology assessment (Li *et al.*, 2014; Wale and Gardner, 2016). Furthermore, over a 5- to 7-day observation period in a TLT system, the total energy dose of the total light exposure time was much lower as compared to light exposure with conventional morphology assessment (Li *et al.*, 2014). In TLT incubators consisting of individual chambers, light exposure is reduced even further. Thus, the use of TLT can standardize variations in light exposure between patients.

Culture environment stability. Compared to conventional embryo assessment, stability of the key environmental parameters may be maintained with TLT (temperature: 0.09–0.2°C; CO₂: 0.1–0.4%; O₂: 0.3–0.5%). Short recovery times for these different parameters are achieved in integrated TLT systems, which are comparable to conventional bench-top incubators. Therefore, TLT provides a safe environment for embryo observation for research and clinical use. Indeed, some studies report that culture in integrated TLT systems may improve embryo development compared to standard incubators (Alhelou *et al.*, 2018; Barberet *et al.*, 2018; Cimadomo *et al.*, 2018; Sciorio *et al.*, 2018; Mascarenhas *et al.*, 2019), while other studies do not confirm this superiority (Cruz *et al.*, 2011; Kirkegaard *et al.*, 2012a; Park *et al.*, 2015; Insua *et al.*, 2017).

Morphokinetic algorithms for embryo selection

Several teams have worked on developing algorithms aimed at standardizing and refining embryo quality evaluation and embryo selection. An algorithm predicting development to the blastocyst stage was first described in 2010, and later validated and adapted (Wong *et al.*, 2010; Meseguer *et al.*, 2011; Conaghan *et al.*, 2013; Rubio *et al.*, 2014; Basile *et al.*, 2015). Although concerns were raised on the reproducibility of the algorithms and cleavage anomalies in previous attempts (Freour *et al.*, 2015; Kirkegaard *et al.*, 2015; Neyer *et al.*, 2015; Barrie *et al.*, 2017), a tendency towards better clinical outcomes was concluded (Petersen *et al.*, 2016; Pribenszky *et al.*, 2017).

Table IV Different commercially available time-lapse systems.

	System A	System B	System C	System D	System E	System F	System G
Incubator	Integrated	Placed in conventional incubators	Integrated	Integrated	Integrated	Integrated	Integrated
External dimensions (W × D × H mm)	530 × 860 × 381	220 × 80 × 110	550 × 600 × 500	600 × 560 × 440	785 × 596 × 380; 960 × 700 × 325	625 × 500 × 300	625 × 500 × 300
Specific culture dish	Single culture	Group culture	Single culture (shared medium)	Single culture	Single or group culture (shared medium)	Group culture (shared medium)	Group culture (shared medium)
Specifications							
Number of focal planes	11 (max.)	3 to 11	11	Up to 17—typically 7	3 to 7	Up to 11	Up to 11
Time between acquisitions	15 min (adjustable between 15 and 60 min)	5 to 60 min	10 min	10 min for 7 focal planes, 2 min for a single focal plane	5 min	5 min	5 min
Camera (megapixels)	1.3	5	2.2 (3 px/μm)	1.3 (3 px/μm)	1.25	5	5
Type of microscopy	Oblique illumination	Brightfield (Hoffman modulation)	Brightfield (Hoffman modulation)	Brightfield (Hoffman Modulation)	brightfield	brightfield	brightfield/darkfield
Embryo illumination for image acquisition	Red LED	Amber LED (590 nm)	Red LED (630 nm)	Red LED (635 nm)	red LED (635 nm)	orange LED (591 nm)	red LED (630 nm)
Time of light exposure	0.008 s	0.2 to 0.005 s	<0.02 s; <32 s/day/embryo	<0.032 s; <31 s/day/embryo (7 focal planes)	0.064 s	<0.005 s; ≈164 s/day/embryo	<0.005 s; ≈203 s/day/embryo
Software							
Morphokinetic annotation	Yes, manual	Yes, manual, guided/semi-automated	Yes, manual, guided/semi-automated	Yes, manual, guided/semi-automated	yes, manual and automated	yes, manual, semi-automated and automated	yes, manual, semi-automated and automated
Predictive algorithm	/	Yes, or defined by user	Yes, or defined by user	Yes, or defined by user	defined by user	defined by user	yes

Continued

Table IV Continued.

	System A	System B	System C	System D	System E	System F	System G
Costs							
General	Culture dish	Culture dish + software	Culture dish + software	Culture dish + software	culture dish + software	culture dish + software	culture dish + software
Gas consumption		N/A	N ₂ : max 5 L/h, typical 2–3 L/h CO ₂ : max 2 L/hr, typical 0.5 L/hr	N ₂ : <10 L/h Typical 3 L/h CO ₂ : <1 L/h	N ₂ : 3–5 L/hr CO ₂ : 1–2 L/hr	N ₂ & CO ₂ : 3.6 L/hr	N ₂ & CO ₂ : 3.6 L/hr
Type of gas	Built-in gas mixer	N/A	Integrated gas mixer	Integrated gas mixer	built in gas mixer; premixed not required	premixed	premixed
Recovery time (min)	Temperature: 10–12 Gas: 5–6	N/A	CO ₂ < 5; O ₂ < 3	CO ₂ and temperature < 5; O ₂ < 15 (with a 30-s door opening, typical door opening is 4 s)	temperature < 1; gas < 3	temperature < 1; CO ₂ < 3; humidity 4 h (for full recovery)	temperature < 1; CO ₂ < 3; humidity 4 h (for full recovery)
Dry or humid culture system	Dry	N/A	Dry	Dry	dry	dry or humid, independently on each chamber	dry or humid, independently on each chamber
pH monitoring	Performed by placing a petri dish with a media sample and oil cover inside the embryo chamber	N/A	Specific pH validation dish	Performed by placing a petri dish with a media sample and oil cover inside the embryo chamber	built in pH measuring system	performed by placing a petri dish with a media sample and oil cover inside the embryo chamber	performed by placing a petri dish with a media sample and oil cover inside the embryo chamber
Capacity	12 embryos/dish; 9 dishes/incubator	16 or 9 embryos/dish; 1 dish/inverted microscope	16 embryos/dish; 15 dishes/incubator	12 embryos/dish; 6 dishes/incubator	14 embryos/dish; 6 or 12 dishes/incubator	16 embryos/dish; 6 dishes/incubator	16 embryos/dish; 6 dishes/incubator
Compartment individualization	Individual sensor for temperature and heating elements; mixed gas provided into each compartment through separated gas line	N/A	Shared chamber	Shared chamber	individual temperature control sensor; one gas mixer supplying all the chambers individually	individual sensors for temperature, humidity and CO ₂ ; individualized heating elements; shared gas supplying all the chambers individually	individual sensors for temperature, humidity and CO ₂ ; individualized heating elements; shared gas supplying all the chambers individually

Continued

Table IV Continued.

	System A	System B	System C	System D	System E	System F	System G
Impact of compartment failure	If one fails in terms of temperature, the rest still works	-	Failure of entire unit (if only the computer systems fails, the incubator is not affected)	Failure of entire unit (if only the computer systems fails, the incubator is not affected)	temperature failure – does not affect the remaining chambers; gas leak - the gas flow is adjusted in the remaining chambers	the damaged compartment can be deactivated.	the damaged compartment can be deactivated.
Impact of camera failure	Incubator works as a regular benchtop	N/A	Incubator works as a regular benchtop	Incubator works as a regular benchtop	incubator works as a regular benchtop	incubator works as a regular benchtop; the remaining non affected cameras work without problem	incubator works as a regular benchtop; the remaining non affected cameras work without problem
Electronic record systems integration	Manually	Possible to integrate	Possible to integrate	Possible to integrate	under development	possible to integrate	possible to integrate
Other	Remote access to the images	Remote access to the images	Culture dishes automatically registered using a barcode labelling; remote access to the images	Remote access to the images	remote access to the images	remote access to the images, dry contact alarm surveillance	remote access to the images, dry contact alarm surveillance

Information was gathered from manufacturers' brochures and through contact with local distributors, from November 2018 until June 2019.

Table V Possible confounding factors with the use of TLT algorithms.

	Parameters to consider	References
Patient-related factors	Age	(Akarsu <i>et al.</i> , 2017, Akhter and Shahab, 2017, Gryshchenko <i>et al.</i> , 2014, Kirkegaard <i>et al.</i> , 2016, Siristatidis <i>et al.</i> , 2015)
	Type of infertility	(Freis <i>et al.</i> , 2018, Sundvall <i>et al.</i> , 2015, Wissing <i>et al.</i> , 2014)
	Weight/BMI/obesity	(Bellver <i>et al.</i> , 2013, Kirkegaard <i>et al.</i> , 2016, Leary <i>et al.</i> , 2015)
	Ovarian stimulation protocol	(Gryshchenko <i>et al.</i> , 2014, Gurbuz <i>et al.</i> , 2016, Kirkegaard <i>et al.</i> , 2016, Munoz <i>et al.</i> , 2013, Wdowiak and Bojar, 2015)
	Type of responder/ovarian reserve	(Akarsu <i>et al.</i> , 2017, Bhide <i>et al.</i> , 2017, Hojnik <i>et al.</i> , 2016, Rienzi <i>et al.</i> , 2015)
	Smoking	(Freour <i>et al.</i> , 2013, Salvarci <i>et al.</i> , 2017, Siristatidis <i>et al.</i> , 2015)
Gamete, embryo or laboratory-related factors	Sperm factor	(Desai <i>et al.</i> , 2018a, Knez <i>et al.</i> , 2013, Lammers <i>et al.</i> , 2015, Mangoli <i>et al.</i> , 2018, Neyer <i>et al.</i> , 2015, Wdowiak <i>et al.</i> , 2015)
	Oocyte morphology	(Mizobe <i>et al.</i> , 2016b, Otsuki <i>et al.</i> , 2018, Van Blerkom, 1990)
	IVM	(Dal Canto <i>et al.</i> , 2016, Eschrich <i>et al.</i> , 2012, Roesner <i>et al.</i> , 2017, Walls <i>et al.</i> , 2015, Wilken-Jensen <i>et al.</i> , 2014)
	Fertilisation technique	(Bodri <i>et al.</i> , 2015, Cruz <i>et al.</i> , 2013, Inoue <i>et al.</i> , 2019, Joergensen <i>et al.</i> , 2014, Kim <i>et al.</i> , 2017, Kirkegaard <i>et al.</i> , 2013b, Kirkegaard <i>et al.</i> , 2013c, Kirkegaard <i>et al.</i> , 2016, Liu <i>et al.</i> , 2015)
	Biopsy	(Bar-El <i>et al.</i> , 2016, Kalma <i>et al.</i> , 2018, Kirkegaard <i>et al.</i> , 2012b)
	Cryopreservation	(Chamayou <i>et al.</i> , 2015, Cobo <i>et al.</i> , 2017, Coello <i>et al.</i> , 2017, De Munck <i>et al.</i> , 2015, Eastick <i>et al.</i> , 2017, Ebner <i>et al.</i> , 2017, Kovacic <i>et al.</i> , 2018, Maezawa <i>et al.</i> , 2014)
	Sex of the embryo	(Bodri <i>et al.</i> , 2016a, Bronet <i>et al.</i> , 2015, Huang <i>et al.</i> , 2019, Serdarogullari <i>et al.</i> , 2014, Zeyad <i>et al.</i> , 2018)

If possible, each laboratory introducing TLT should perform a proper validation, based on appropriate sample size or post-hoc power analysis, certifying the value of each variable introduced and the corrections for putative confounders that could influence the algorithms (Table V) (Carrasco *et al.*, 2017).

Evidence of a clinical benefit of TLT

Like any new intervention, TLT should be implemented in routine clinical practice only after stringent tests demonstrating a benefit for patients (Brison *et al.*, 2013; Harper *et al.*, 2017). However, a clear increase in IVF success rates with the use of TLT remains to be proven.

The latest Cochrane review (nine RCTs, 2955 women) (Armstrong *et al.*, 2019) reported insufficient evidence for differences in live birth rate (odds ratio (OR) 1.12, 95% CI 0.92–1.36), miscarriage rate (OR 0.63, 95% CI 0.45–0.89) or clinical pregnancy rate (OR 0.95, 95% CI 0.78–1.16) for TLT combined with embryo selection software versus conventional incubation and assessment. Likewise, a putative benefit of TLT was not demonstrated by meta-analyses (Polanski *et al.*, 2014; Armstrong *et al.*, 2015a; Racowsky *et al.*, 2015; Chen *et al.*, 2017; Armstrong *et al.*, 2018). Conversely, one meta-analysis, using a different methodological approach, has suggested a beneficial effect of TLT compared to conventional incubation and assessment, respectively, reporting a significantly higher ongoing pregnancy rate (51.0 versus 39.9%; OR 1.54, 95% CI 1.21–1.97), a significantly lower early pregnancy loss (15.3 versus 21.3%; OR 0.66, 95% CI 0.47–0.94) and a significantly increased live birth rate (44.2 versus 31.3%; OR 1.67, 95% CI 1.13–2.46) (Pribenszky *et al.*, 2017).

Cumulative live birth rates were assessed in a recent retrospective study of 1882 cycles comparing time-lapse and conventional incubation/assessment (Mascarenhas *et al.*, 2019): the study showed similar cumulative live birth rates for time-lapse and conventional incubation/assessment (51.7 versus 51.2%; OR 1.02, 95% CI: 0.85–1.22), although fresh embryo transfer live birth rates were higher for TLT cycles (36.8 versus 33.9%, adjusted OR 1.28, 95% CI: 1.05–1.57).

The main reason for the controversy over TLT efficacy is the fact that it entails two distinct components, i.e. an undisturbed incubation environment and embryo selection through imaging software. In this respect, these two components have not been effectively distinguished in the majority of studies, possibly masking the weight of the effect of better culture conditions or improved embryo selection on the reported outcomes (Armstrong *et al.*, 2015b). Additional confounders that may explain the heterogeneity amongst studies include different days of assessment, different endpoints, the wide array of morphokinetic timings assessed, inter- and intra-operator variation in annotating and the various other confounding factors listed in Table V.

Importantly, no safety issues have been reported following embryo culture in TLT incubators, and obstetric and perinatal outcomes, such as duration of gestation, congenital malformations and birth weight, are comparable (Costa-Borges *et al.*, 2016; Insua *et al.*, 2017; Kovacs *et al.*, 2019) or better (Mascarenhas *et al.*, 2019) compared to standard incubation.

Despite the current lack of evidence from RCTs for a clinical benefit of TLT, it is reasonable to assume that, compared with static observations, continuous embryo monitoring in an undisturbed environment will offer more information into embryo development and is expected to enhance the identification of good-prognosis embryos for clinical

use. In order to firmly establish a putative beneficial effect of TLT, more well-designed and sufficiently powered RCTs reporting on live births and perinatal outcomes are necessary.

Current state of TLT

Although in-house systems have existed since the late 1990s, TLT became commercially available for human IVF in 2009. The large volume of published articles, communications in congresses and active communication on the internet and in conventional media from IVF centres using TLT suggest a vigorous implementation rate of this technology in IVF laboratories throughout the world. Surprisingly, almost no data are currently available on the global use of TLT. Scotland represents a somewhat unique area, since the Scottish government funding has enabled all four publicly funded (UK National Health Service) assisted conception units within the country to invest in TLT (Thomas Freour, personal communication). Besides this specific case, only two surveys could be found reporting on TLT implementation rate and use. The first study by Dolinko *et al.* reported the results of an online survey of 294 IVF laboratory directors in the USA on TLT use (Dolinko *et al.*, 2017). In total, 162 (55%) responded, with 35 laboratories (17%) reporting that they run at least one TLT system. The presence and availability of TLT was positively associated with the number of IVF cycles performed in the centre. Following this first report, a French team conducted a very similar survey of 210 laboratory directors in all 105 IVF laboratories in France (Boueilh *et al.*, 2018). Amongst the 78 respondents (response rate 37%), 30 (39%) reported using TLT clinically. Amongst non-users, 11 (23%) reported plans to invest in TLT within the next 2 years. Unlike the situation in the USA, TLT implementation was not associated with the number of IVF cycles performed in France. Although these two studies provide an interesting insight into TLT implementation in two different countries in terms of IVF regulation or funding policy, it is not prudent to draw a conclusion of the overall use of TLT worldwide. Altogether, these elements highlight the relevance of the present guidelines in order to help patients and clinics benefit as much as possible from TLT. A more global picture of the TLT market would be interesting in order to evaluate its current use and trends in IVF practice and to find opportunities for cost-effectiveness and medical studies.

Current and future research perspectives

In comparison with the rapid technical development of TLT together with other technologies for basic research in cell biology, the TLT in clinical embryology remains in its infancy and, as such, there is significant scope to refine and improve the method. However, beyond this, the type of data generated, coupled with the relative ease of use and non-invasive nature of TLT, means that there are exciting prospects for exploring fundamental developmental biology in significant detail.

Embryo selection parameters based on visual indicators of presumed quality have largely been a subjective application of a decision tree (Simopoulou *et al.*, 2018). The inclusion of multiple visual parameters has led to improvements in outcomes, and the widespread application of the so-called 'Gardner criteria' is a good illustrative example (Gardner and Schoolcraft, 1999). This indicates the prospective value of assessing multiple parameters, and data generated by TLT will

offer the opportunity for profound evolution of such multiparameter analyses.

Artificial intelligence (AI), or machine learning, describes a non-biased approach to multiparameter analysis. In the context of TLT, attempts are underway to use higher-powered computer-processing power to analyse large data sets of images to identify combinations of parameters that might link to embryo viability. There is little doubt that the future of AI and TLT will incorporate some degree of machine learning, to facilitate complex analysis of large data sets, which will likely reveal currently unidentified combinations of visual markers. Khosravi and colleagues used AI and TLT and, by analysing more than 10 000 embryos, developed a model that was able to predict blastocyst quality with an AUC of >0.98 (Khosravi *et al.*, 2019). Using a similar approach, Tran and colleagues have recently reported the development of a deep learning model to annotate automatically morphokinetic videos. The authors retrospectively analysed more than 10 000 videos from multiple centres and were able to show that their model was able to reproducibly identify images from blastocysts that went on to give a foetal heartbeat, with an AUC of >0.90 (Tran *et al.*, 2019).

An important issue that deserves discussion is whether biological justification is required for acceptance of computer-generated algorithms to select embryos based on machine-learned combinations of parameters. The very strength of adopting an objective approach of using AI to interrogate digital images free of human bias is that such a system will 'look beyond' traditional parameters of morphology and may identify unique combinations of markers that relate to embryo viability. However, in doing so, it is possible that such combinations may be unfamiliar. Furthermore, as AI systems are not able to ascribe meaning to parameters, it is possible that markers may relate to non-classical features, such as image grey scale or image texture depth (e.g. Molder *et al.*, 2015). Before adoption of such approaches, there is a requirement for robust clinical validation prior to evaluating its acceptance by the relevant stakeholders.

TLT enables research possibilities in fundamental developmental processes. For example, the immediate period after fertilisation is characterized by a number of molecular processes, each of them with its own specific dynamics. With TLT, it has been possible to observe a number of processes in their entirety including cytoplasmic movements in oocytes during meiosis resumption (Bui *et al.*, 2017) and in embryos (Milewski *et al.*, 2018), fertilisation events (Coticchio *et al.*, 2018), the beginning of first mitosis (Wong *et al.*, 2010) and the dynamics of blastocyst formation (Marcos *et al.*, 2015). The observation of such crucial developmental events in real time has revealed a number of new parameters that have been introduced into embryology (Table III). Moreover, with a more detailed understanding of developmental kinetics, we may be able to ascribe key landmarks to other aspects of embryo physiology, such as embryo chromosomal status (Pennetta *et al.*, 2018) and response to cryopreservation (Taborin and Kovacic, 2019).

Looking forward, it is difficult to imagine that there will not be significant improvements to the technology of TLT to drive further knowledge and understanding of early development. Such developments are likely to come from more refined image collection methods and the integration with other technologies. Development of fluorescence and confocal time-lapse imaging enables the observation of morphokinetics of organelles and chromosomes during oocyte maturation (Duncan *et al.*, 2012; Holubcova *et al.*, 2015; Patel *et al.*, 2015; Zielinska *et al.*,

2015; Capalbo *et al.*, 2017). Furthermore, exciting developments in fluorescence live-cell imaging of human embryos (Hashimoto *et al.*, 2016), fluorescent light sheet in toto imaging of developing mouse embryos (Strnad *et al.*, 2016) and a combination of the fluorescent time-lapse with comparative genomic hybridisation (Chavez *et al.*, 2012) or single-cell sequencing (Daughtry *et al.*, 2019) of individual blastomeres may help in discovering mechanisms of aneuploidy during cleavages of primate and human embryos.

There is growing interest in using advanced label-free imaging techniques to gain a molecular-level understanding of cellular function (Kaspruwicz *et al.*, 2017). Such approaches can yield additional information on the physiology of the cell, including details of metabolic processes, since many metabolites and enzymes exhibit autofluorescent properties (Gosnell *et al.*, 2016). Measuring metabolic and biochemical function has long been a pursuit of those interested in the identification of biomarkers of viability. Bradley *et al.* (2016) have used an image-based approach called coherent anti-stokes raman scattering to identify reliably the composition, ratio and real-time change in lipid profiles of single preimplantation embryos (Bradley *et al.*, 2016). Sutton-McDowall *et al.* (2017) were able to demonstrate differences in metabolic profiles of embryos grown in hyperoxic (20%) or normoxic (7%) conditions using hyperspectral imaging to measure ratios of NAD(P)H and flavin adenine dinucleotide (Sutton-McDowall *et al.*, 2017). Similarly, Sanchez *et al.* (2018) used fluorescent lifetime imaging (to detect mitochondrial dysfunction (Sanchez *et al.*, 2018).

The search for objective robust biomarkers of embryo viability continues, although to date, and despite significant research effort, no single reliable biomarker with a sufficiently high predictability of live birth has yet been identified: this may reflect the complexity of preimplantation development. Consequently, the search for biomarkers must no longer occur in isolation; the combination of TLT with other markers of embryo physiology is a natural evolution of both fields.

Should TLT data be shared with patients

The introduction of TLT in assisted reproduction has raised many questions and concerns, mostly related to its clinical relevance in IVF and its impact on reproductive outcome. However, the sustained implementation of this technology has raised many additional logistic questions associated with daily practice. Some of the most pressing concerns are how we engage with the final stakeholders, our patients.

For example, we may consider whether we should declare the brand of the TLT in the reports provided to our patients, but is there any need to link the information provided with product used to obtain such information? Since many TLT incubators are technically similar and able to provide comparable results, it may be prudent to avoid speaking in terms of 'brands'.

TLT practitioners may also wish to consider the number of images used in describing the embryo(s) selected for transfer. Ideally, we may choose three images for the embryos when day 3 transfer is performed and four for blastocyst (day 5–6) transfers. Additionally, accurate time-stamping of images is crucial; for example, fertilisation (18 h), day 2 (44 h) and day 3 (68 h), with the intention to accurately describe the development of the embryo up to blastocyst stage (116 h), as described in the Istanbul Consensus (Alpha Scientists in Reproductive Medicine and ESHRE Special Interest Group of Embryology,

2011). Important technical difficulties appear when patients present high numbers of oocytes to be fertilized, more than those that can be placed in a single slide; this may necessitate the need for several TLT slides per patient, each being considered as a 'new' or 'different' patient for the TLT software. This could result in reports being not immediately compatible and cause confusion for doctors and even patients.

In the report, together with the images of the embryos transferred (at different times) thought needs to be given as to whether to include images (at least one per embryo) of those to be vitrified, or those to be discarded. In such images, the time reference (after ICSI) or the stage of embryo development may be included as headings. The presentation of the embryos, to distinguish those embryos to be transferred from those frozen or discarded, could be potentially useful. We may also consider including multiple images at different time frames of embryos that are vitrified. The amount of information that may be used for a report could be debated, too many pages may create confusion and too little may result in a deficient information.

Inclusion of single static representative images does not address how to share data on morphokinetics or morphology with patients. We may add information about the timings of key landmarks in embryo development, as well as the incidence of abnormal or irregular cleavages, blastocyst collapses or multinucleation, as potential parameters that may affect negatively the implantation potential. Together with this, there remains the option to share the classification of the embryos after using any of the algorithms described in the scientific literature methods of embryo development. However, the inclusion of such complex information means it may be very difficult for patients to fully understand it at first mention, or that it will need extra time with the patient at consultation to explain those values.

The obligation of the medical professionals should be to inform patients objectively about the development phase of the implementation of new technology in clinical practice. Thus, the question remains what clinicians should tell their patients. We need to explain that we do not have perfect tools to identify the best embryo, but we may change the order in which the embryos are transferred based on these technologies, which may not improve the cumulative outcome by itself, but may impact time to pregnancy (Kovacs and Lieman, 2019). The additional financial expenses should be taken into consideration and also the most suitable indication, which is still unknown. However, it is wise to explain that TLT still lacks a convincing evidence base to prove any clinical efficacy, although it may provide otherwise unknown information on embryo quality and development. In addition, TLT may help to counsel couples in decisions making regarding further treatment, donor egg use, adoption, etc.

There are few publications that provide scientific evidence of possible benefits of sharing TLT data with the patients. Blomqvist *et al.* explored patient-oriented aspects of new technologies by a prospective, observational questionnaire with a sample size of more than 200 patients. Interestingly, the majority of them found that viewing the videos and obtaining a copy of it relevant, but only if the treatment resulted in a viable pregnancy (Blomqvist *et al.*, 2017). Reinforcing the importance of the videos, Bui *et al.* reported the relevance of remote access to images of developing embryos during an IVF cycle. In their study of over 100 patients, the majority surveyed indicated that viewing their embryo images during the cycle enhanced their experience of IVF treatment (Bui *et al.* 2018).

Currently, the working group considers that 'TLT are next generation incubators that allow a detailed real-time embryo evaluation. The continuous embryo monitoring facilitates a complete follow up and a detailed analysis of embryo development. With TLT it is possible to perform a study of the kinetics of embryo development and the relationship between timings of cleavages and embryo viability. This information may therefore help to identify good embryos and recognize those with numerous atypical embryo developmental patterns. However, it should be noted that in these development stages there is an extraordinary plasticity in embryo morphology and developmental dynamics and that embryos also have their own self-correction mechanisms. With more research morphokinetics in the future will improve its power as an adjunctive test to select embryos with the highest implantation potential/deselect embryos with lowest implantation potential'.

Conclusion

TLT has been introduced into human IVF as a routine procedure only in the last decade, much later than in other fields of biosciences, and yet it has led to major changes in the way that embryos are observed and handled. When TLT was first adopted, expectations were high. It was hoped that dynamic observation of development would offer a more precise, non-invasive modality to assess embryo viability, with obvious implications for the efficiency of ART treatment. Many studies, although mainly retrospective, have attempted to answer the question of whether TLT brings a clinical benefit, without reaching a consensus. The hopes are not lost, but thus far, studies to effectively assess the efficacy of TLT for embryo selection have lacked sufficient rigour to demonstrate unequivocally any substantial improvement in pregnancy rate/live birth outcomes. Regardless of a possible direct impact on clinical outcome, TLT does confer several advantages that justify its use. Its introduction in the workflow of the IVF laboratory, however, has a multiplicity of implications requiring technical and managerial expertise, as well as a strategic vision for this technology. This manuscript has attempted to collate recommendations to assist with the choice, introduction, management and harnessing of the TLT in the IVF laboratory.

Based on current technology, TLT probably offers the safest and most stable embryo culture environment. Continued embryo monitoring has allowed us to identify previously unknown or undetectable aspects of development, some of which, such as direct cleavage of the fertilized egg into three blastomeres, have significant clinical impact. There is now awareness that chromosomal aberrations may affect embryo morphokinetics, but not to an extent to suggest that TLT can replace PGT-A in the identification of euploid embryos. TLT devices, however, are relatively demanding pieces of equipment. Therefore, a suitable technical choice requires elements of knowledge relevant to embryo culture conditions, consistency of use between operators and laboratories, data management, cost-benefit balance and its potential for research. Making patients aware of the benefits and limits of TLT is not simple, but every effort should be made to inform them in a meaningful and unbiased fashion. The promise that TLT may evolve into a full-blown embryo selection modality, once combined with AI and non-invasive analytical approaches, is compelling. While the prediction of future achievements of TLT is a difficult exercise, there is little doubt that this technology is here to stay. Mastering its use is therefore becoming imperative for embryologists and IVF laboratories.

Supplementary data

Supplementary data are available at [Human Reproduction Open](https://academic.oup.com/hropen/article-abstract/2020/2/hoaa008/5609428) online.

Acknowledgements

The working group also thanks the reviewers that participated in the stakeholder review for their critical and constructive remarks.

Authors' roles

All authors contributed equally in discussing the recommendations until consensus. N.L.C. provided methodological and organisational support.

Funding

The meetings of the working group were funded by the European Society of Human Reproduction and Embryology.

Conflict of interest

S.A. declares participation in the Nordic Embryology Academic Team with meetings sponsored by Gedeon Richter. T.E. declares to have organized workshops for Esco and receiving consulting fees from Merck and Gynemed and speakers' fees from Merck and Esco. T.F. received consulting fees from Vitrolife and Laboratoires Génévrier, and speakers' fees from Merck Serono, Gedeon Richter, MSD, Laboratoires Génévrier and Ferring. M.M. reports non-financial support from Merck, SA, outside the submitted work. M.M.E. received speakers' fees from Merck, Ferring and MSD. R.S. received a research grant from ESHRE. G.C. received speakers' fees from IBSA and Excemed. The other authors declare that they have no conflict of interest.

References

- Adolfsson E, Andershed AN. Morphology vs morphokinetics: a retrospective comparison of inter-observer and intra-observer agreement between embryologists on blastocysts with known implantation outcome. *JBRA Assist. Reprod.* 2018;**22**:228–237.
- Aguilar J, Motato Y, Escriba MJ, Ojeda M, Munoz E, Meseguer M. The human first cell cycle: impact on implantation. *Reprod Biomed Online* 2014;**28**:475–484.
- Ahlstrom A, Westin C, Reismer E, Wikland M, Hardarson T. Trophoctoderm morphology: an important parameter for predicting live birth after single blastocyst transfer. *Hum Reprod* 2011;**26**:3289–3296.
- Akarsu S, Gode F, Isik AZ, Celenk H, Tamer FB, Erkilinc S. Comparison of the morphokinetic parameters of embryos according to ovarian reserve in IVF cycles. *Gynecol Endocrinology* 2017;**33**:733–736.
- Akhter N, Shahab M. Morphokinetic analysis of human embryo development and its relationship to the female age: a retrospective time-lapse imaging study. *Cell Mol Biol (Noisy-le-Grand)* 2017;**63**:84–92.
- Alhelou Y, Mat Adenan NA, Ali J. Embryo culture conditions are significantly improved during uninterrupted incubation: a randomized controlled trial. *Reprod Biol* 2018;**18**:40–45.

- Alpha Scientists in Reproductive Medicine and ESHRE Special Interest Group of Embryology. The Istanbul consensus workshop on embryo assessment: proceedings of an expert meeting[†]. *Hum Reprod* 2011;**26**:1270–1283.
- Armstrong S, Arroll N, Cree LM, Jordan V, Farquhar C. Time-lapse systems for embryo incubation and assessment in assisted reproduction. *Cochrane Database Syst Rev* 2015a;Cd011320.
- Armstrong S, Bhide P, Jordan V, Pacey A, Farquhar C. Time-lapse systems for embryo incubation and assessment in assisted reproduction. *Cochrane Database Syst Rev* 2018;**5**:Cd011320.
- Armstrong S, Bhide P, Jordan V, Pacey A, Marjoribanks J, Farquhar C. Time-lapse systems for embryo incubation and assessment in assisted reproduction. *Cochrane Database Syst Rev* 2019;Cd011320.
- Armstrong S, Vail A, Mastenbroek S, Jordan V, Farquhar C. Time-lapse in the IVF-lab: how should we assess potential benefit? *Hum Reprod* 2015b;**30**:3–8.
- Athayde Wirka K, Chen AA, Conaghan J, Ivani K, Gvakharia M, Behr B, Suraj V, Tan L, Shen S. Atypical embryo phenotypes identified by time-lapse microscopy: high prevalence and association with embryo development. *Fertil Steril* 2014;**101**:1637–1648.e1631–e1635.
- Azzarello A, Hoest T, Mikkelsen AL. The impact of pronuclei morphology and dynamicity on live birth outcome after time-lapse culture. *Hum Reprod* 2012;**27**:2649–2657.
- Balakier H, Sojecki A, Motamedi G, Librach C. Impact of multinucleated blastomeres on embryo developmental competence, morphokinetics, and aneuploidy. *Fertil Steril* 2016;**106**:608, e602–614.
- Baltz JM. Media composition: salts and osmolality. *Methods in molecular biology (Clifton, NJ)* 2012;**912**:61–80.
- Baltz JM, Tartia AP. Cell volume regulation in oocytes and early embryos: connecting physiology to successful culture media. *Hum Reprod Update* 2010;**16**:166–176.
- Bar-El L, Kalma Y, Malcov M, Schwartz T, Raviv S, Cohen T, Amir H, Cohen Y, Reches A, Amit A et al. Blastomere biopsy for PGD delays embryo compaction and blastulation: a time-lapse microscopic analysis. *J Assist Reprod Genet* 2016;**33**:1449–1457.
- Barberet J, Chammas J, Bruno C, Valot E, Vuillemin C, Jonval L, Choux C, Sagot P, Soudry A, Fauque P. Randomized controlled trial comparing embryo culture in two incubator systems: G185 K-System versus EmbryoScope. *Fertil Steril* 2018;**109**:302, e301–309.
- Barnes FL, Crombie A, Gardner DK, Kausche A, Lacham-Kaplan O, Suikkari AM, Tiglias J, Wood C, Trounson AO. Blastocyst development and birth after in-vitro maturation of human primary oocytes, intracytoplasmic sperm injection and assisted hatching. *Hum Reprod* 1995;**10**:3243–3247.
- Barrie A, Homburg R, McDowell G, Brown J, Kingsland C, Troup S. Examining the efficacy of six published time-lapse imaging embryo selection algorithms to predict implantation to demonstrate the need for the development of specific, in-house morphokinetic selection algorithms. *Fertil Steril* 2017;**107**:613–621.
- Basile N, Morbeck D, Garcia-Velasco J, Bronet F, Meseguer M. Type of culture media does not affect embryo kinetics: a time-lapse analysis of sibling oocytes. *Hum Reprod* 2013;**28**:634–641.
- Basile N, Vime P, Florensa M, Aparicio Ruiz B, Garcia Velasco JA, Remohi J, Meseguer M. The use of morphokinetics as a predictor of implantation: a multicentric study to define and validate an algorithm for embryo selection. *Hum Reprod* 2015;**30**:276–283.
- Bellver J, Mifsud A, Grau N, Privitera L, Meseguer M. Similar morphokinetic patterns in embryos derived from obese and normoweight infertile women: a time-lapse study. *Hum Reprod* 2013;**28**:794–800.
- Bhide P, Escriba M, Srikantharajah A, Joshi H, Gudi A, Shah A, Acharya G, Homburg R. Anti-Mullerian hormone (AMH) and embryo quality assessed by time-lapse imaging (TLI): a cross-sectional observational study. *Arch Gynecol Obstet* 2017;**296**:583–587.
- Blomqvist AB, Samir R, Engström A, Nilsson S, Lind A, Westlander G, Larsson S, Hardarson T, Hreinsson J. Patients experience of viewing time-lapse sequences: a prospective survey study. *Gynecol Reproduct Endocrinol-UK* 2017;1–6.
- Bodri D, Kawachiya S, Sugimoto T, Yao Serna J, Kato R, Matsumoto T. Time-lapse variables and embryo gender: a retrospective analysis of 81 live births obtained following minimal stimulation and single embryo transfer. *J Assist Reprod Genet* 2016a;**33**:589–596.
- Bodri D, Sugimoto T, Serna JY, Kondo M, Kato R, Kawachiya S, Matsumoto T. Influence of different oocyte insemination techniques on early and late morphokinetic parameters: retrospective analysis of 500 time-lapse monitored blastocysts. *Fertil Steril* 2015;**104**:1175–1181.e1171–1172.
- Bodri D, Sugimoto T, Yao Serna J, Kawachiya S, Kato R, Matsumoto T. Blastocyst collapse is not an independent predictor of reduced live birth: a time-lapse study. *Fertil Steril* 2016b;**105**:1476, e1473–1483.
- Bontekoe S, Mantikou E, van Wely M, Seshadri S, Repping S, Mastenbroek S. Low oxygen concentrations for embryo culture in assisted reproductive technologies. *Cochrane Database Syst Rev* 2012;Cd008950.
- Boueilh T, Reignier A, Barriere P, Freour T. Time-lapse imaging systems in IVF laboratories: a French national survey. *J Assist Reprod Genet* 2018;**35**:2181–2186.
- Bradley J, Pope I, Masia F, Sanusi R, Langbein W, Swann K, Borri P. Quantitative imaging of lipids in live mouse oocytes and early embryos using CARS microscopy. *Development* 2016;**143**:2238–2247.
- Brison DR, Roberts SA, Kimber SJ. How should we assess the safety of IVF technologies? *Reprod Biomed Online* 2013;**27**:710–721.
- Bronet F, Nogales MC, Martinez E, Ariza M, Rubio C, Garcia-Velasco JA, Meseguer M. Is there a relationship between time-lapse parameters and embryo sex? *Fertil Steril* 2015;**103**:396, e392–401.
- Bui TTH, Belli M, Fassina L, Vigone G, Merico V, Garagna S, Zuccotti M. Cytoplasmic movement profiles of mouse surrounding nucleolus and not-surrounding nucleolus antral oocytes during meiotic resumption. *Mol Reprod Dev* 2017;**84**:356–362.
- Bui D, Debenham E, Hesketh N, McArthur S, Waite K & Bowman M. Remote access to embryo images and video during their cycle enhances a patients In Vitro Fertilisation (IVF) experience. *Human Reproduction* 2018;**33**:Issue suppl_1, i403.
- Campbell A, Fishel S, Bowman N, Duffy S, Sedler M, Hickman CF. Modelling a risk classification of aneuploidy in human embryos using non-invasive morphokinetics. *Reprod Biomed Online* 2013;**26**:477–485.
- Capalbo A, Hoffmann ER, Cimadomo D, Ubaldi FM, Rienzi L. Human female meiosis revised: new insights into the mechanisms of chromosome segregation and aneuploidies from advanced genomics and time-lapse imaging. *Hum Reprod Update* 2017;**23**:706–722.
- Carrasco B, Arroyo G, Gil Y, Gomez MJ, Rodriguez I, Barri PN, Veiga A, Boada M. Selecting embryos with the highest implantation potential

- using data mining and decision tree based on classical embryo morphology and morphokinetics. *J Assist Reprod Genet* 2017;**34**:983–990.
- Cetinkaya M, Pirkevi C, Yelke H, Colakoglu YK, Atayurt Z, Kahraman S. Relative kinetic expressions defining cleavage synchronicity are better predictors of blastocyst formation and quality than absolute time points. *J Assist Reprod Genet* 2015;**32**:27–35.
- Chamayou S, Patrizio P, Storaci G, Tomaselli V, Alecci C, Ragolia C, Crescenzo C, Guglielmino A. The use of morphokinetic parameters to select all embryos with full capacity to implant. *J Assist Reprod Genet* 2013;**30**:703–710.
- Chamayou S, Romano S, Alecci C, Storaci G, Ragolia C, Palagiano A, Guglielmino A. Oocyte vitrification modifies nucleolar remodeling and zygote kinetics—a sibling study. *J Assist Reprod Genet* 2015;**32**:581–586.
- Chavez SL, Loewke KE, Han J, Moussavi F, Colls P, Munne S, Behr B, Reijo Pera RA. Dynamic blastomere behaviour reflects human embryo ploidy by the four-cell stage. *Nat Commun* 2012;**3**:1251.
- Chen M, Wei S, Hu J, Yuan J, Liu F. Does time-lapse imaging have favorable results for embryo incubation and selection compared with conventional methods in clinical in vitro fertilization? A meta-analysis and systematic review of randomized controlled trials. *PLoS One* 2017;**12**:e0178720.
- Chung YH, Hsiao YH, Kao WL, Hsu CH, Yao DJ, Chen C. Microwells support high-resolution time-lapse imaging and development of preimplanted mouse embryos. *Biomicrofluidics* 2015;**9**:022407.
- Cimadomo D, Scarica C, Maggiulli R, Orlando G, Soscia D, Albricci L, Romano S, Sanges F, Ubaldi FM, Rienzi L. Continuous embryo culture elicits higher blastulation but similar cumulative delivery rates than sequential: a large prospective study. *J Assist Reprod Genet* 2018;**35**:1329–1338.
- Ciray HN, Aksoy T, Goktas C, Ozturk B, Bahceci M. Time-lapse evaluation of human embryo development in single versus sequential culture media—a sibling oocyte study. *J Assist Reprod Genet* 2012;**29**:891–900.
- Ciray HN, Campbell A, Agerholm IE, Aguilar J, Chamayou S, Esbert M, Sayed S. Proposed guidelines on the nomenclature and annotation of dynamic human embryo monitoring by a time-lapse user group. *Hum Reprod* 2014;**29**:2650–2660.
- Cobo A, Coello A, Remohi J, Serrano J, de Los Santos JM, Meseguer M. Effect of oocyte vitrification on embryo quality: time-lapse analysis and morphokinetic evaluation. *Fertil Steril* 2017;**108**:491, e493–497.
- Coello A, Meseguer M, Galan A, Alegre L, Remohi J, Cobo A. Analysis of the morphological dynamics of blastocysts after vitrification/warming: defining new predictive variables of implantation. *Fertil Steril* 2017;**108**:659, e654–666.
- Conaghan J, Chen AA, Willman SP, Ivani K, Chenette PE, Boostanfar R, Baker VL, Adamson GD, Abusief ME, Gvakharia M et al. Improving embryo selection using a computer-automated time-lapse image analysis test plus day 3 morphology: results from a prospective multicenter trial. *Fertil Steril* 2013;**100**:412, e415–419.
- Costa-Borges N, Belles M, Meseguer M, Galliano D, Ballesteros A, Calderon G. Blastocyst development in single medium with or without renewal on day 3: a prospective cohort study on sibling donor oocytes in a time-lapse incubator. *Fertil Steril* 2016;**105**:707–713.
- Coticchio G, Lagalla C, Sturmey R, Pennetta F, Borini A. The enigmatic morula: mechanisms of development, cell fate determination, self-correction and implications for ART. *Hum Reprod Update* 2019;**25**:422–438.
- Coticchio G, Mignini Renzini M, Novara PV, Lain M, De Ponti E, Turchi D, Fadini R, Dal CM. Focused time-lapse analysis reveals novel aspects of human fertilization and suggests new parameters of embryo viability. *Hum Reprod* 2018;**33**:23–31.
- Cruz M, Gadea B, Garrido N, Pedersen KS, Martinez M, Perez-Cano I, Munoz M, Meseguer M. Embryo quality, blastocyst and ongoing pregnancy rates in oocyte donation patients whose embryos were monitored by time-lapse imaging. *J Assist Reprod Genet* 2011;**28**:569–573.
- Cruz M, Garrido N, Gadea B, Munoz M, Perez-Cano I, Meseguer M. Oocyte insemination techniques are related to alterations of embryo developmental timing in an oocyte donation model. *Reprod Biomed Online* 2013;**27**:367–375.
- Cruz M, Garrido N, Herrero J, Perez-Cano I, Munoz M, Meseguer M. Timing of cell division in human cleavage-stage embryos is linked with blastocyst formation and quality. *Reprod Biomed Online* 2012;**25**:371–381.
- Cummins JM, Breen TM, Harrison KL, Shaw JM, Wilson LM, Hennessey JF. A formula for scoring human embryo growth rates in in vitro fertilization: its value in predicting pregnancy and in comparison with visual estimates of embryo quality. *J In Vitro Fert Embryo Transf* 1986;**3**:284–295.
- Dahdouh EM, Balayla J, Audibert F, Wilson RD, Audibert F, Brock JA, Campagnolo C, Carroll J, Chong K, Gagnon A et al. Technical update: Preimplantation genetic diagnosis and screening. *Journal d'obstetrique et gynecologie du Canada* 2015;**37**:451–463.
- Dal Canto M, Coticchio G, Mignini Renzini M, De Ponti E, Novara PV, Brambillasca F, Comi R, Fadini R. Cleavage kinetics analysis of human embryos predicts development to blastocyst and implantation. *Reprod Biomed Online* 2012;**25**:474–480.
- Dal Canto M, Novara PV, Coticchio G, Mignini Renzini M, Brambillasca F, Brigante C, De Ponti E, Fadini R. Morphokinetics of embryos developed from oocytes matured in vitro. *J Assist Reprod Genet* 2016;**33**:247–253.
- Daughtry BL, Rosenkrantz JL, Lazar NH, Fei SS, Redmayne N, Torkenczy KA, Adey A, Yan M, Gao L, Park B et al. Single-cell sequencing of primate preimplantation embryos reveals chromosome elimination via cellular fragmentation and blastomere exclusion. *Genome Res* 2019;**29**:367–382.
- De los Santos MJ, Apter S, Coticchio G, Debrock S, Lundin K, Plancha CE, Prados F, Rienzi L, Verheyen G, Woodward B et al. revised guidelines for good practice in IVF laboratories (2015). *Hum Reprod* 2016;**31**:685–686.
- De Munck N, Petrusa L, Verheyen G, Staessen C, Vandekelde Y, Sterckx J, Bocken G, Jacobs K, Stoop D, De Rycke M et al. Chromosomal meiotic segregation, embryonic developmental kinetics and DNA (hydroxy)methylation analysis consolidate the safety of human oocyte vitrification. *Mol Hum Reprod* 2015;**21**:535–544.
- Desai N, Gill P, Tadros NN, Goldberg JM, Sabanegh E, Falcone T. Azoospermia and embryo morphokinetics: testicular sperm-derived embryos exhibit delays in early cell cycle events and increased arrest prior to compaction. *J Assist Reprod Genet* 2018a;**35**:1339–1348.
- Desai N, Goldberg JM, Austin C, Falcone T. Are cleavage anomalies, multinucleation, or specific cell cycle kinetics observed with time-lapse imaging predictive of embryo developmental capacity or ploidy? *Fertil Steril* 2018b;**109**:665–674.

- Desai N, Ploskonka S, Goodman LR, Austin C, Goldberg J, Falcone T. Analysis of embryo morphokinetics, multinucleation and cleavage anomalies using continuous time-lapse monitoring in blastocyst transfer cycles. *Reprod Biol Endocrinol* 2014;**12**:54.
- Dolinko AV, Farland LV, Kaser DJ, Missmer SA, Racowsky C. National survey on use of time-lapse imaging systems in IVF laboratories. *J Assist Reprod Genet* 2017;**34**:1167–1172.
- Duncan FE, Hornick JE, Lampson MA, Schultz RM, Shea LD, Woodruff TK. Chromosome cohesion decreases in human eggs with advanced maternal age. *Aging Cell* 2012;**11**:1121–1124.
- Eastick J, Venetis C, Cooke S, Storr A, Susetio D, Chapman M. Is early embryo development as observed by time-lapse microscopy dependent on whether fresh or frozen sperm was used for ICSI? A cohort study. *J Assist Reprod Genet* 2017;**34**:733–740.
- Ebner T, Oppelt P, Radler E, Allerstorfer C, Habelsberger A, Mayer RB, Shebl O. Morphokinetics of vitrified and warmed blastocysts predicts implantation potential. *J Assist Reprod Genet* 2017;**34**:239–244.
- Ebner T, Shebl O, Moser M, Mayer RB, Arzt W, Tews G. Group culture of human zygotes is superior to individual culture in terms of blastulation, implantation and live birth. *Reprod Biomed Online* 2010;**21**:762–768.
- Ergin EG, Caliskan E, Yalcinkaya E, Oztel Z, Cokelez K, Ozay A, Ozornek HM. Frequency of embryo multinucleation detected by time-lapse system and its impact on pregnancy outcome. *Fertil Steril* 2014;**102**:1029, e1021–1033.
- Escrich L, Grau N, de los Santos MJ, Romero JL, Pellicer A, Escriba MJ. The dynamics of in vitro maturation of germinal vesicle oocytes. *Fertil Steril* 2012;**98**:1147–1151.
- Ezoe K, Ohata K, Morita H, Ueno S, Miki T, Okimura T, Uchiyama K, Yabuuchi A, Kobayashi T, Montag M et al. Prolonged blastomere movement induced by the delay of pronuclear fading and first cell division adversely affects pregnancy outcomes after fresh embryo transfer on day 2: a time-lapse study. *Reprod Biomed Online* 2019;**38**:659–668.
- Fan YL, Han SB, Wu LH, Wang YP, Huang GN. Abnormally cleaving embryos are able to produce live births: a time-lapse study. *J Assist Reprod Genet* 2016;**33**:379–385.
- Fawzy M, AbdelRahman MY, Zidan MH, Abdel Hafez FF, Abdelghafar H, Al-Inany H, Bedaiwy MA. Humid versus dry incubator: a prospective, randomized, controlled trial. *Fertil Steril* 2017;**108**:277–283.
- Ferrick L, Lee YSL, Gardner DK. Reducing time to pregnancy and facilitating the birth of healthy children through functional analysis of embryo physiology. *Biol Reprod* 2019;**101**:1124–1139.
- Fischer B, Bavister BD. Oxygen tension in the oviduct and uterus of rhesus monkeys, hamsters and rabbits. *J Reprod Fertil* 1993;**99**:673–679.
- Fishel S, Campbell A, Montgomery S, Smith R, Nice L, Duffy S, Jenner L, Berrisford K, Kellam L, Smith R et al. Time-lapse imaging algorithms rank human preimplantation embryos according to the probability of live birth. *Reprod Biomed Online* 2018;**37**:304–313.
- Freis A, Dietrich JE, Binder M, Holschbach V, Strowitzki T, Germeyer A. Relative morphokinetics assessed by time-lapse imaging are altered in embryos from patients with endometriosis. *Reprod Sci* 2018;**25**:1279–1285.
- Freour T, Dessolle L, Lammers J, Lattes S, Barriere P. Comparison of embryo morphokinetics after in vitro fertilization-intracytoplasmic sperm injection in smoking and nonsmoking women. *Fertil Steril* 2013;**99**:1944–1950.
- Freour T, Le Fleuter N, Lammers J, Spingart C, Reignier A, Barriere P. External validation of a time-lapse prediction model. *Fertil Steril* 2015;**103**:917–922.
- Fulka H, Kyogoku H, Zatssepina O, Langerova A, Fulka J Jr. Can nucleoli be markers of developmental potential in human zygotes? *Trends Mol Med* 2015;**21**:663–672.
- Gardner DK, Schoolcraft B. In vitro culture of human blastocyst. In: Jansen R, Mortimer D (eds). *Towards Reproductive Certainty: Fertility and Genetics Beyond 1999*. UK: Parthenon Publishing Carnforth, 1999, 378–388.
- Goodman LR, Goldberg J, Falcone T, Austin C, Desai N. Does the addition of time-lapse morphokinetics in the selection of embryos for transfer improve pregnancy rates? A randomized controlled trial. *Fertil Steril* 2016;**105**:275–285.e210.
- Gopichandran N, Leese HJ. The effect of paracrine/autocrine interactions on the in vitro culture of bovine preimplantation embryos. *Reproduction* 2006;**131**:269–277.
- Gosnell ME, Anwer AG, Mahbub SB, Menon Perinchery S, Inglis DW, Adhikary PP, Jazayeri JA, Cahill MA, Saad S, Pollock CA et al. Quantitative non-invasive cell characterisation and discrimination based on multispectral autofluorescence features. *Sci Rep* 2016;**6**:23453.
- Griffin DK, Ogur C. Chromosomal analysis in IVF: just how useful is it? *Reproduction* 2018;**156**:F29–F50.
- Gryshchenko MG, Pravdyuk AI, Parashchuk VY. Analysis of factors influencing morphokinetic characteristics of embryos in ART cycles. *Gynecol Endocrinol* 2014;**30**:6–8.
- Guerif F, Le Gouge A, Giraudeau B, Poindron J, Bidault R, Gasnier O, Royere D. Limited value of morphological assessment at days 1 and 2 to predict blastocyst development potential: a prospective study based on 4042 embryos. *Hum Reprod* 2007;**22**:1973–1981.
- Gurbuz AS, Gode F, Uzman MS, Ince B, Kaya M, Ozcimen N, Ozcimen EE, Acar A. GnRH agonist triggering affects the kinetics of embryo development: a comparative study. *J Ovarian Research* 2016;**9**:22.
- Hardarson T, Bungum M, Conaghan J, Meintjes M, Chantilis SJ, Molnar L, Gunnarsson K, Wikland M. Noninferiority, randomized, controlled trial comparing embryo development using media developed for sequential or undisturbed culture in a time-lapse setup. *Fertil Steril* 2015;**104**:1452–1459, e1451–1454.
- Hardarson T, Lofman C, Coull G, Sjogren A, Hamberger L, Edwards RG. Internalization of cellular fragments in a human embryo: time-lapse recordings. *Reprod Biomed Online* 2002;**5**:36–38.
- Harper J, Jackson E, Sermon K, Aitken RJ, Harbottle S, Mocanu E, Hardarson T, Mathur R, Viville S, Vail A et al. Adjuncts in the IVF laboratory: where is the evidence for 'add-on' interventions? *Hum Reprod* 2017;**32**:485–491.
- Hashimoto S, Nakano T, Yamagata K, Inoue M, Morimoto Y, Nakaoka Y. Multinucleation per se is not always sufficient as a marker of abnormality to decide against transferring human embryos. *Fertil Steril* 2016;**106**:133–139.e136.
- Heo YS, Cabrera LM, Song JW, Futai N, Tung YC, Smith GD, Takayama S. Characterization and resolution of evaporation-mediated osmolality shifts that constrain microfluidic cell culture in poly(dimethylsiloxane) devices. *Anal Chem* 2007;**79**:1126–1134.

- Hojnik N, Vlaisavljevic V, Kovacic B. Morphokinetic characteristics and developmental potential of in vitro cultured embryos from natural cycles in patients with poor ovarian response. *Biomed Res Int* 2016;**2016**:4286528.
- Holubcova Z, Blayney M, Elder K, Schuh M. Human oocytes. Error-prone chromosome-mediated spindle assembly favors chromosome segregation defects in human oocytes. *Science* 2015;**348**:1143–1147.
- Huang B, Ren X, Zhu L, Wu L, Tan H, Guo N, Wei Y, Hu J, Liu Q, Chen W *et al.* Is differences in embryo morphokinetic development significantly associated with human embryo sex? *Biol Reprod* 2019;**100**:618–623.
- Inoue T, Uemura M, Miyazaki K, Yamashita Y. Failure of complete hatching of ICSI-derived human blastocyst by cell herniation via small slit and insufficient expansion despite ongoing cell proliferation. *J Assist Reprod Genet* 2019;**36**:1579–1589.
- Insua MF, Cobo AC, Larreategui Z, Ferrando M, Serra V, Meseguer M. Obstetric and perinatal outcomes of pregnancies conceived with embryos cultured in a time-lapse monitoring system. *Fertil Steril* 2017;**108**:498–504.
- Iwata K, Yumoto K, Sugishima M, Mizoguchi C, Kai Y, Iba Y, Mio Y. Analysis of compaction initiation in human embryos by using time-lapse cinematography. *J Assist Reprod Genet* 2014;**31**:421–426.
- Joergensen MW, Agerholm I, Hindkjaer J, Bolund L, Sunde L, Ingerslev HJ, Kirkegaard K. Altered cleavage patterns in human tripronuclear embryos and their association to fertilization method: a time-lapse study. *J Assist Reprod Genet* 2014;**31**:435–442.
- Kalma Y, Bar-El L, Asaf-Tisser S, Malcov M, Reches A, Hasson J, Amir H, Azem F, Ben-Yosef D. Optimal timing for blastomere biopsy of 8-cell embryos for preimplantation genetic diagnosis. *Hum Reprod* 2018;**33**:32–38.
- Kaser DJ, Racowsky C. Clinical outcomes following selection of human preimplantation embryos with time-lapse monitoring: a systematic review. *Hum Reprod Update* 2014;**20**:617–631.
- Kaspruwicz R, Suman R, O'Toole P. Characterising live cell behaviour: traditional label-free and quantitative phase imaging approaches. *Int J Biochem Cell Biol* 2017;**84**:89–95.
- Kazdar N, Brugnon F, Bouche C, Jouve G, Veau S, Drapier H, Rousseau C, Pimentel C, Viard P, Belaud-Rotureau MA *et al.* Comparison of human embryomorphokinetic parameters in sequential or global culture media. *Ann Biol Clin* 2017;**75**:403–410.
- Kelley RL, Gardner DK. Combined effects of individual culture and atmospheric oxygen on preimplantation mouse embryos in vitro. *Reprod Biomed Online* 2016;**33**:537–549.
- Kelley RL, Gardner DK. In vitro culture of individual mouse preimplantation embryos: the role of embryo density, microwells, oxygen, timing and conditioned media. *Reprod Biomed Online* 2017;**34**:441–454.
- Khosravi P, Kazemi E, Zhan Q, Malmsten JE, Toschi M, Zisimopoulos P, Sigras A, Lavery S, Cooper LAD, Hickman C *et al.* Deep learning enables robust assessment and selection of human blastocysts after in vitro fertilization. *NPJ Digital Med* 2019;**2**:21.
- Kim HJ, Yoon HJ, Jang JM, Lee WD, Yoon SH, Lim JH. Evaluation of human embryo development in in vitro fertilization- and intracytoplasmic sperm injection-fertilized oocytes: a time-lapse study. *Clin Exp Reprod Med* 2017;**44**:90–95.
- Kirkegaard K, Ahlstrom A, Ingerslev HJ, Hardarson T. Choosing the best embryo by time lapse versus standard morphology. *Fertil Steril* 2015;**103**:323–332.
- Kirkegaard K, Hindkjaer JJ, Grondahl ML, Kesmodel US, Ingerslev HJ. A randomized clinical trial comparing embryo culture in a conventional incubator with a time-lapse incubator. *J Assist Reprod Genet* 2012a;**29**:565–572.
- Kirkegaard K, Hindkjaer JJ, Ingerslev HJ. Human embryonic development after blastomere removal: a time-lapse analysis. *Hum Reprod* 2012b;**27**:97–105.
- Kirkegaard K, Hindkjaer JJ, Ingerslev HJ. Effect of oxygen concentration on human embryo development evaluated by time-lapse monitoring. *Fertil Steril* 2013a;**99**:738–744.e734.
- Kirkegaard K, Hindkjaer JJ, Ingerslev HJ. Hatching of in vitro fertilized human embryos is influenced by fertilization method. *Fertil Steril* 2013b;**100**:1277–1282.
- Kirkegaard K, Kesmodel US, Hindkjaer JJ, Ingerslev HJ. Time-lapse parameters as predictors of blastocyst development and pregnancy outcome in embryos from good prognosis patients: a prospective cohort study. *Hum Reprod* 2013c;**28**:2643–2651.
- Kirkegaard K, Sundvall L, Erlandsen M, Hindkjaer JJ, Knudsen UB, Ingerslev HJ. Timing of human preimplantation embryonic development is confounded by embryo origin. *Hum Reprod* 2016;**31**:324–331.
- Knez K, Tomazevic T, Vrtacnik-Bokal E, Virant-Klun I. Developmental dynamics of IMSI-derived embryos: a time-lapse prospective study. *Reprod Biomed Online* 2013;**27**:161–171.
- Kovacic B, Sajko MC, Vlaisavljevic V. A prospective, randomized trial on the effect of atmospheric versus reduced oxygen concentration on the outcome of intracytoplasmic sperm injection cycles. *Fertil Steril* 2010;**94**:511–519.
- Kovacic B, Taborin M, Vlaisavljevic V. Artificial blastocoel collapse of human blastocysts before vitrification and its effect on re-expansion after warming - a prospective observational study using time-lapse microscopy. *Reprod Biomed Online* 2018;**36**:121–129.
- Kovacs P, Lieman HJ. Which embryo selection method should be offered to the patients? *J Assist Reprod Genet* 2019;**36**:603–605.
- Kovacs P, Matyas S, Forgacs V, Sajko A, Molnar L, Pribenszky C. Non-invasive embryo evaluation and selection using time-lapse monitoring: results of a randomized controlled study. *Eur J Obstet Gynecol Reprod Biol* 2019;**233**:58–63.
- Lagalla C, Tarozzi N, Sciajno R, Wells D, Di Santo M, Nadalini M, Distratis V, Borini A. Embryos with morphokinetic abnormalities may develop into euploid blastocysts. *Reprod Biomed Online* 2017;**34**:137–146.
- Lammers J, Reigner A, Spingart C, Catteau A, David L, Barriere P, Freour T. Does sperm origin affect embryo morphokinetic parameters? *J Assist Reprod Genet* 2015;**32**:1325–1332.
- Le Cruguel S, Ferre-L'Hotellier V, Moriniere C, Lemerle S, Reynier P, Descamps P, May-Panloup P. Early compaction at day 3 may be a useful additional criterion for embryo transfer. *J Assist Reprod Genet* 2013;**30**:683–690.
- Leary C, Leese HJ, Sturmey RG. Human embryos from overweight and obese women display phenotypic and metabolic abnormalities. *Hum Reprod* 2015;**30**:122–132.
- Lee E, Costello MF, Botha WC, Illingworth P, Chambers GM. A cost-effectiveness analysis of preimplantation genetic testing for aneuploidy (PGT-A) for up to three complete assisted reproductive technology cycles in women of advanced maternal age. *Aust N Z J Obstet Gynaecol* 2019;**59**:573–579.

- Lemmen JG, Agerholm I, Ziebe S. Kinetic markers of human embryo quality using time-lapse recordings of IVF/ICSI-fertilized oocytes. *Reprod Biomed Online* 2008;**17**:385–391.
- Li R, Pedersen KS, Liu Y, Pedersen HS, Laegdsmand M, Rickelt LF, Kuhl M, Callesen H. Effect of red light on the development and quality of mammalian embryos. *J Assist Reprod Genet* 2014;**31**:795–801.
- Li W, Goossens K, Van Poucke M, Forier K, Braeckmans K, Van Soom A, Peelmann LJ. High oxygen tension increases global methylation in bovine 4-cell embryos and blastocysts but does not affect general retrotransposon expression. *Reprod Fertil Dev* 2016;**28**:948–959.
- Liu Y, Chapple V, Feenan K, Roberts P, Matson P. Time-lapse videography of human embryos: using pronuclear fading rather than insemination in IVF and ICSI cycles removes inconsistencies in time to reach early cleavage milestones. *Reprod Biol* 2015;**15**:122–125.
- Liu Y, Chapple V, Roberts P, Ali J, Matson P. Time-lapse videography of human oocytes following intracytoplasmic sperm injection: events up to the first cleavage division. *Reprod Biol* 2014a;**14**:249–256.
- Liu Y, Chapple V, Roberts P, Matson P. Prevalence, consequence, and significance of reverse cleavage by human embryos viewed with the use of the Embryoscope time-lapse video system. *Fertil Steril* 2014b;**102**:1295–1300.e1292.
- Maezawa T, Yamanaka M, Hashimoto S, Amo A, Ohgaki A, Nakaoka Y, Fukuda A, Ikeda T, Inoue M, Morimoto Y. Possible selection of viable human blastocysts after vitrification by monitoring morphological changes. *J Assist Reprod Genet* 2014;**31**:1099–1104.
- Mangoli E, Khalili MA, Talebi AR, Ghasemi-Esmailabad S, Hosseini A. Is there any correlation between sperm parameters and chromatin quality with embryo morphokinetics in patients with male infertility? *Andrologia* 2018;**50**:e12997.
- Marcos J, Perez-Albala S, Mifsud A, Molla M, Landeras J, Meseguer M. Collapse of blastocysts is strongly related to lower implantation success: a time-lapse study. *Hum Reprod* 2015;**30**:2501–2508.
- Martinez-Granados L, Serrano M, Gonzalez-Utor A, Ortiz N, Badajoz V, Olaya E, Prados N, Boada M, Castilla JA. Inter-laboratory agreement on embryo classification and clinical decision: Conventional morphological assessment vs time lapse. *PLoS One* 2017;**12**:e0183328.
- Mascarenhas M, Fox SJ, Thompson K, Balen AH. Cumulative live birth rates and perinatal outcomes with the use of time-lapse imaging incubators for embryo culture: a retrospective cohort study of 1882 ART cycles. *BJOG* 2019;**126**:280–286.
- Meintjes M, Chantilis SJ, Douglas JD, Rodriguez AJ, Guerami AR, Bookout DM, Barnett BD, Madden JD. A controlled randomized trial evaluating the effect of lowered incubator oxygen tension on live births in a predominantly blastocyst transfer program. *Hum Reprod* 2009;**24**:300–307.
- Meseguer M, Herrero J, Tejera A, Hilligsoe KM, Ramsing NB, Remohi J. The use of morphokinetics as a predictor of embryo implantation. *Hum Reprod* 2011;**26**:2658–2671.
- Meseguer M, Rubio I, Cruz M, Basile N, Marcos J, Requena A. Embryo incubation and selection in a time-lapse monitoring system improves pregnancy outcome compared with a standard incubator: a retrospective cohort study. *Fertil Steril* 2012;**98**:1481–1489.e1410.
- Milewski R, Szpila M, Ajduk A. Dynamics of cytoplasm and cleavage divisions correlates with preimplantation embryo development. *Reproduction* 2018;**155**:1–14.
- Mio Y, Iwata K, Yumoto K, Maeda K. Human Embryonic Behavior Observed with Time-Lapse Cinematography. *J Health Med Informat* 2014;**5**:143.
- Mizobe Y, Ezono Y, Tokunaga M, Oya N, Iwakiri R, Yoshida N, Sato Y, Onoue N, Miyoshi K. Selection of human blastocysts with a high implantation potential based on timely compaction. *J Assist Reprod Genet* 2017;**34**:991–997.
- Mizobe Y, Oya N, Iwakiri R, Yoshida N, Sato Y, Miyoshi K, Tokunaga M, Ezono Y. Effects of early cleavage patterns of human embryos on subsequent in vitro development and implantation. *Fertil Steril* 2016a;**106**:348–353.e342.
- Mizobe Y, Oya N, Iwakiri R, Yoshida N, Sato Y, Onoue N, Miyoshi K, Tokunaga M, Ezono Y. Developmental ability of embryos produced from oocytes with fragile oolemma by intracytoplasmic sperm injection. *J Assist Reprod Genet* 2016b;**33**:1685–1690.
- Mizobe Y, Tokunaga M, Oya N, Iwakiri R, Yoshida N, Sato Y, Onoue N, Ezono Y. Synchrony of the first division as an index of the blastocyst formation rate during embryonic development. *Reprod Medicine Biol* 2018;**17**:64–70.
- Molder A, Drury S, Costen N, Hartshorne GM, Czanner S. Semiautomated analysis of embryoscope images: using localized variance of image intensity to detect embryo developmental stages. *Cytometry A* 2015;**87**:119–128.
- Motato Y, de los Santos MJ, Escriba MJ, Ruiz BA, Remohi J, Meseguer M. Morphokinetic analysis and embryonic prediction for blastocyst formation through an integrated time-lapse system. *Fertil Steril* 2016;**105**:376–384.e379.
- Munoz M, Cruz M, Humaidan P, Garrido N, Perez-Cano I, Meseguer M. The type of GnRH analogue used during controlled ovarian stimulation influences early embryo developmental kinetics: a time-lapse study. *Eur J Obstet Gynecol Reprod Biol* 2013;**168**:167–172.
- Nakahara T, Iwase A, Goto M, Harata T, Suzuki M, Ienaga M, Kobayashi H, Takikawa S, Manabe S, Kikkawa F, Ando H. Evaluation of the safety of time-lapse observations for human embryos. *J Assist Reprod Genet* 2010;**27**:93–96.
- Neal SA, Morin SJ, Franasiak JM, Goodman LR, Juneau CR, Forman EJ, Werner MD, Scott RT Jr. Preimplantation genetic testing for aneuploidy is cost-effective, shortens treatment time, and reduces the risk of failed embryo transfer and clinical miscarriage. *Fertil Steril* 2018;**110**:896–904.
- Neyer A, Zintz M, Stecher A, Bach M, Wirleitner B, Zech NH, Vanderzwalmen P. The impact of paternal factors on cleavage stage and blastocyst development analyzed by time-lapse imaging—a retrospective observational study. *J Assist Reprod Genet* 2015;**32**:1607–1614.
- O'Neill C. The potential roles for embryotrophic ligands in preimplantation embryo development. *Hum Reprod Update* 2008;**14**:275–288.
- Oh SJ, Gong SP, Lee ST, Lee EJ, Lim JM. Light intensity and wavelength during embryo manipulation are important factors for maintaining viability of preimplantation embryos in vitro. *Fertil Steril* 2007;**88**:1150–1157.
- Otsuki J, Iwasaki T, Katada Y, Tsutsumi Y, Tsuji Y, Furuhashi K, Koikeguchi S, Shiotani M. A higher incidence of cleavage failure in oocytes containing smooth endoplasmic reticulum clusters. *J Assist Reprod Genet* 2018;**35**:899–905.
- Otsuki J, Iwasaki T, Tsuji Y, Katada Y, Sato H, Tsutsumi Y, Hatano K, Furuhashi K, Matsumoto Y, Koikeguchi S et al. Potential of zygotes

- to produce live births can be identified by the size of the male and female pronuclei just before their membranes break down. *Reprod Med Biol* 2017;**16**:200–205.
- Ottosen LD, Hindkjaer J, Ingerslev J. Light exposure of the ovum and preimplantation embryo during ART procedures. *J Assist Reprod Genet* 2007;**24**:99–103.
- Park H, Bergh C, Selleskog U, Thurin-Kjellberg A, Lundin K. No benefit of culturing embryos in a closed system compared with a conventional incubator in terms of number of good quality embryos: results from an RCT. *Hum Reprod* 2015;**30**:268–275.
- Patel J, Tan SL, Hartshorne GM, McAinsh AD. Unique geometry of sister kinetochores in human oocytes during meiosis I may explain maternal age-associated increases in chromosomal abnormalities. *Biol Open* 2015;**5**:178–184.
- Payne D, Flaherty SP, Barry MF, Matthews CD. Preliminary observations on polar body extrusion and pronuclear formation in human oocytes using time-lapse video cinematography. *Hum Reprod* 1997;**12**:532–541.
- Pennetta F, Lagalla C, Borini A. Embryo morphokinetic characteristics and euploidy. *Curr Opin Obstet Gynecol* 2018;**30**:185–196.
- Penzias A, Bendikson K, Butts S, Coutifaris C, Falcone T, Fossom G, Gitlin S, Gracia C, Hansen K, La Barbera A et al. The use of preimplantation genetic testing for aneuploidy (PGT-A): a committee opinion. *Fertil Steril* 2018;**109**:429–436.
- Petersen BM, Boel M, Montag M, Gardner DK. Development of a generally applicable morphokinetic algorithm capable of predicting the implantation potential of embryos transferred on day 3. *Hum Reprod* 2016;**31**:2231–2244.
- Polanski LT, Coelho Neto MA, Nastri CO, Navarro PA, Ferriani RA, Raine-Fenning N, Martins WP. Time-lapse embryo imaging for improving reproductive outcomes: systematic review and meta-analysis. *Ultrasound Obstet Gynecol* 2014;**44**:394–401.
- Pribenszky C, Nilselid AM, Montag M. Time-lapse culture with morphokinetic embryo selection improves pregnancy and live birth chances and reduces early pregnancy loss: a meta-analysis. *Reprod Biomed Online* 2017;**35**:511–520.
- Racowsky C, Kovacs P, Martins WP. A critical appraisal of time-lapse imaging for embryo selection: where are we and where do we need to go? *J Assist Reprod Genet* 2015;**32**:1025–1030.
- Reignier A, Lammers J, Barriere P, Freour T. Can time-lapse parameters predict embryo ploidy? A systematic review. *Reprod Biomed Online* 2018;**36**:380–387.
- Rienzi L, Capalbo A, Stoppa M, Romano S, Maggiulli R, Albricci L, Scarica C, Farcomeni A, Vajta G, Ubaldi FM. No evidence of association between blastocyst aneuploidy and morphokinetic assessment in a selected population of poor-prognosis patients: a longitudinal cohort study. *Reprod Biomed Online* 2015;**30**:57–66.
- Rijnders PM, Jansen CA. The predictive value of day 3 embryo morphology regarding blastocyst formation, pregnancy and implantation rate after day 5 transfer following in-vitro fertilization or intracytoplasmic sperm injection. *Hum Reprod* 1998;**13**:2869–2873.
- Rinaudo PF, Giritharan G, Talbi S, Dobson AT, Schultz RM. Effects of oxygen tension on gene expression in preimplantation mouse embryos. *Fertil Steril* 2006;**86**:1252–1265, 1265.e1251–1236.
- Rocafort E, Enciso M, Leza A, Sarasa J, Aizpurua J. Euploid embryos selected by an automated time-lapse system have superior SET outcomes than selected solely by conventional morphology assessment. *J Assist Reprod Genet* 2018;**35**:1573–1583.
- Roesner S, Dietrich JE, Weigert J, Montag M, Toth B, Strowitzki T. Time-lapse imaging reveals differences in growth dynamics of embryos after in vitro maturation compared with conventional stimulation. *Fertil Steril* 2017;**107**:606–612.e603.
- Rosenwaks Z, Handyside AH, Fiorentino F, Gleicher N, Paulson RJ, Schattman GL, Scott RT Jr, Summers MC, Treff NR, Xu K. The pros and cons of preimplantation genetic testing for aneuploidy: clinical and laboratory perspectives. *Fertil Steril* 2018;**110**:353–361.
- Rubio I, Galan A, Larreategui Z, Ayerdi F, Bellver J, Herrero J, Meseguer M. Clinical validation of embryo culture and selection by morphokinetic analysis: a randomized, controlled trial of the EmbryoScope. *Fertil Steril* 2014;**102**:1287–1294.e1285.
- Rubio I, Kuhlmann R, Agerholm I, Kirk J, Herrero J, Escriba MJ, Bellver J, Meseguer M. Limited implantation success of direct-cleaved human zygotes: a time-lapse study. *Fertil Steril* 2012;**98**:1458–1463.
- Salvarci A, Gurbuz AS, Uzman S, Kaya M, Gorkemli H. Comparison of embryo morphokinetics following intracytoplasmic sperm injection in smoker and non-smoker couples: are the results different? *J Pak Med Assoc* 2017;**67**:1552–1557.
- Sanchez T, Wang T, Pedro MV, Zhang M, Esencan E, Sakkas D, Needleman D, Seli E. Metabolic imaging with the use of fluorescence lifetime imaging microscopy (FLIM) accurately detects mitochondrial dysfunction in mouse oocytes. *Fertil Steril* 2018;**110**:1387–1397.
- Schiewe MC, Whitney JB, Nugent N, Zozula S, Anderson RE. Human blastocyst development: A randomized comparison of sibling zygotes cultured in Vitrolife G-TL™ to Life Global® single-step media. *J Pregnancy Reprod* 2018; Open Access text.
- Sciorio R, Thong JK, Pickering SJ. Comparison of the development of human embryos cultured in either an EmbryoScope or benchtop incubator. *J Assist Reprod Genet* 2018;**35**:515–522.
- Scott L. The biological basis of non-invasive strategies for selection of human oocytes and embryos. *Hum Reprod Update* 2003;**9**:237–249.
- Sepulveda S, Garcia J, Arriaga E, Diaz J, Noriega-Portella L, Noriega-Hoces L. In vitro development and pregnancy outcomes for human embryos cultured in either a single medium or in a sequential media system. *Fertil Steril* 2009;**91**:1765–1770.
- Serdarogullari M, Findikli N, Goktas C, Sahin O, Ulug U, Yagmur E, Bahceci M. Comparison of gender-specific human embryo development characteristics by time-lapse technology. *Reprod Biomed Online* 2014;**29**:193–199.
- Sermon K, Capalbo A, Cohen J, Coonen E, De Rycke M, De Vos A, Delhanty J, Fiorentino F, Gleicher N, Griesinger G et al. The why, the how and the when of PGS 2.0: current practices and expert opinions of fertility specialists, molecular biologists, and embryologists. *Mol Hum Reprod* 2016;**22**:845–857.
- Sfontouris IA, Martins WP, Nastri CO, Viana IG, Navarro PA, Raine-Fenning N, van der Poel S, Rienzi L, Racowsky C. Blastocyst culture using single versus sequential media in clinical IVF: a systematic review and meta-analysis of randomized controlled trials. *J Assist Reprod Genet* 2016;**33**:1261–1272.
- Simopoulou M, Sfakianoudis K, Maziotis E, Antoniou N, Rapani A, Anifandis G, Bakas P, Bolaris S, Pantou A, Pantos K et al. Are computational applications the “crystal ball” in the IVF laboratory?

- The evolution from mathematics to artificial intelligence. *J Assist Reprod Genet* 2018;**35**:1545–1557.
- Siristatidis C, Komitopoulou MA, Makris A, Sialakouma A, Botzaki M, Mastorakos G, Salamalekis G, Bettocchi S, Palmer GA. Morphokinetic parameters of early embryo development via time lapse monitoring and their effect on embryo selection and ICSI outcomes: a prospective cohort study. *J Assist Reprod Genet* 2015;**32**:563–570.
- Somigliana E, Busnelli A, Paffoni A, Vigano P, Riccaboni A, Rubio C, Capalbo A. Cost-effectiveness of preimplantation genetic testing for aneuploidies. *Fertil Steril* 2019;**111**:1169–1176.
- Storr A, Venetis C, Cooke S, Kilani S, Ledger W. Time-lapse algorithms and morphological selection of day-5 embryos for transfer: a preclinical validation study. *Fertil Steril* 2018;**109**:276–283.e273.
- Storr A, Venetis CA, Cooke S, Susetio D, Kilani S, Ledger W. Morphokinetic parameters using time-lapse technology and day 5 embryo quality: a prospective cohort study. *J Assist Reprod Genet* 2015;**32**:1151–1160.
- Strnad P, Gunther S, Reichmann J, Krzic U, Balazs B, de Medeiros G, Norlin N, Hiragi T, Hufnagel L, Ellenberg J. Inverted light-sheet microscope for imaging mouse pre-implantation development. *Nat Methods* 2016;**13**:139–142.
- Summers MC, Bhatnagar PR, Lawitts JA, Biggers JD. Fertilization in vitro of mouse ova from inbred and outbred strains: complete preimplantation embryo development in glucose-supplemented KSOM. *Biol Reprod* 1995;**53**:431–437.
- Sunde A, Brison D, Dumoulin J, Harper J, Lundin K, Magli MC, Van den Abbeel E, Veiga A. Time to take human embryo culture seriously. *Hum Reprod* 2016;**31**:2174–2182.
- Sundvall L, Ingerslev HJ, Breth Knudsen U, Kirkegaard K. Inter- and intra-observer variability of time-lapse annotations. *Hum Reprod* 2013;**28**:3215–3221.
- Sundvall L, Kirkegaard K, Ingerslev HJ, Knudsen UB. Unaltered timing of embryo development in women with polycystic ovarian syndrome (PCOS): a time-lapse study. *J Assist Reprod Genet* 2015;**32**:1031–1042.
- Sutton-McDowall ML, Gosnell M, Anwer AG, White M, Purdey M, Abell AD, Goldys EM, Thompson JG. Hyperspectral microscopy can detect metabolic heterogeneity within bovine post-compaction embryos incubated under two oxygen concentrations (7% versus 20%). *Hum Reprod* 2017;**32**:2016–2025.
- Swain JE, Cabrera L, Xu X, Smith GD. Microdrop preparation factors influence culture-media osmolality, which can impair mouse embryo preimplantation development. *Reprod Biomed Online* 2012;**24**:142–147.
- Tabarin M, Kovacic B. Morphometric protocol for the objective assessment of blastocyst behavior during vitrification and warming steps. *J Vis Exp* 2019. e58540.
- Tran D, Cooke S, Illingworth PJ, Gardner DK. Deep learning as a predictive tool for fetal heart pregnancy following time-lapse incubation and blastocyst transfer. *Hum Reprod* 2019;**34**:1011–1018.
- Umaoka Y, Noda Y, Nakayama T, Narimoto K, Mori T, Iritani A. Effect of visual light on in vitro embryonic development in the hamster. *Theriogenology* 1992;**38**:1043–1054.
- Van Blerkom J. Occurrence and developmental consequences of aberrant cellular organization in meiotically mature human oocytes after exogenous ovarian hyperstimulation. *J Electron Microsc Tech* 1990;**16**:324–346.
- Vermeulen N, Le Clef N, Veleza Z, D'Angelo A, Tilleman K. European Recommendations for good practice in addition to an evidence-based guidelines programme: rationale and method of development. *BMJ Evidence-Based Medicine*. 2019; **24**:30–34.
- Vinals Gonzalez X, Odia R, Cawood S, Gaunt M, Saab W, Seshadri S, Serhal P. Contraction behaviour reduces embryo competence in high-quality euploid blastocysts. *J Assist Reprod Genet* 2018;**35**:1509–1517.
- Wale PL, Gardner DK. Time-lapse analysis of mouse embryo development in oxygen gradients. *Reprod Biomed Online* 2010;**21**:402–410.
- Wale PL, Gardner DK. Oxygen regulates amino acid turnover and carbohydrate uptake during the preimplantation period of mouse embryo development. *Biol Reprod* 2012;**87**:21–28.
- Wale PL, Gardner DK. The effects of chemical and physical factors on mammalian embryo culture and their importance for the practice of assisted human reproduction. *Hum Reprod Update* 2016; **22**:2–22.
- Walls ML, Ryan JP, Keelan JA, Hart R. In vitro maturation is associated with increased early embryo arrest without impairing morphokinetic development of useable embryos progressing to blastocysts. *Hum Reprod* 2015;**30**:1842–1849.
- Wdowiak A, Bakalczuk S, Bakalczuk G. The effect of sperm DNA fragmentation on the dynamics of the embryonic development in intracytoplasmic sperm injection. *Reprod Biol* 2015;**15**:94–100.
- Wdowiak A, Bojar I. Ovarian stimulation with human and recombinant gonadotropin - comparison of in vitro fertilization efficiency with use of time-lapse monitoring. *Reprod Health* 2015;**12**:113.
- Werner MD, Hong KH, Franasiak JM, Forman EJ, Reda CV, Molinaro TA, Upham KM, Scott RT Jr. Sequential versus Monophasic Media Impact Trial (SuMMIT): a paired randomized controlled trial comparing a sequential media system to a monophasic medium. *Fertil Steril* 2016;**105**:1215–1221.
- Wilken-Jensen HN, Kristensen SG, Jeppesen JV, Yding AC. Developmental competence of oocytes isolated from surplus medulla tissue in connection with cryopreservation of ovarian tissue for fertility preservation. *Acta Obstet Gynecol Scand* 2014;**93**:32–37.
- Wissing ML, Bjerre MR, Olesen AI, Hoest T, Mikkelsen AL. Impact of PCOS on early embryo cleavage kinetics. *Reprod Biomed Online* 2014;**28**:508–514.
- Wolff HS, Fredrickson JR, Walker DL, Morbeck DE. Advances in quality control: mouse embryo morphokinetics are sensitive markers of in vitro stress. *Hum Reprod* 2013;**28**:1776–1782.
- Wong CC, Loewke KE, Bossert NL, Behr B, De Jonge CJ, Baer TM, Reijo Pera RA. Non-invasive imaging of human embryos before embryonic genome activation predicts development to the blastocyst stage. *Nat Biotechnol* 2010;**28**:1115–1121.
- Wu L, Han W, Zhang X, Wang J, Liu W, Xiong S, Huang G. A retrospective analysis of morphokinetic parameters according to the implantation outcome of IVF treatment. *Eur J Obstet Gynecol Reprod Biol* 2016;**197**:186–190.
- Yang HW, Hwang KJ, Kwon HC, Kim HS, Choi KW, Oh KS. Detection of reactive oxygen species (ROS) and apoptosis in human fragmented embryos. *Hum Reprod* 1998;**13**:998–1002.
- Yumoto K, Iwata K, Sugishima M, Yamauchi J, Nakaoka M, Tsuneto M, Shimura T, Flaherty S, Mio Y. Unstable osmolality of microdrops cultured in non-humidified incubators. *J Assist Reprod Genet* 2019;**36**:1571–1577.

- Zaninovic N, Irani M, Meseguer M. Assessment of embryo morphology and developmental dynamics by time-lapse microscopy: is there a relation to implantation and ploidy? *Fertil Steril* 2017;**108**:722–729.
- Zeyad A, Al-Abdulkareem B, Hammadeh ME. The relationship between preimplantation morphokinetics of human embryos and sex chromosome pattern. *Reprod Biol* 2018;**18**:385–389.
- Zhan Q, Ye Z, Clarke R, Rosenwaks Z, Zaninovic N. Direct unequal cleavages: embryo developmental competence, genetic constitution and clinical outcome. *PLoS One* 2016;**11**:e0166398.
- Zielinska AP, Holubcova Z, Blayney M, Elder K, Schuh M. Sister kinetochore splitting and precocious disintegration of bivalents could explain the maternal age effect. *eLife* 2015;**4**:e11389.